

7470A


LB 93157

INSTRUMENT ID: CV1

Linear

 μ Abs.:

36625



A= 0.0000e+000

B= 2.7443e-004

C= -1.1191e-002

Rho= 0.9999070

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.012	-0.012	-3	0.000	-3				
0.2	0.200	0.208	0.008	798	0.0 %	798				
2.5	2.500	2.498	-0.002	9143	0.0 %	9143				
5.0	5.000	5.063	0.063	18491	0.0 %	18491				
7.5	7.500	7.403	-0.097	27017	0.0 %	27017				
10.0	10.000	10.040	0.040	36625	0.0 %	36625				

LB93157 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Feb 2018 09:12:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	-3	-	0.0000	7470A	PPB	08 Feb 2018 09:15:23	S	Std
0.2 - 1	S0.2	798	-	0.2000	7470A	PPB	08 Feb 2018 09:17:31	S	Std
2.5 - 1	S2.5	9143	-	2.5000	7470A	PPB	08 Feb 2018 09:19:40	S	Std
5.0 - 1	S5	18491	-	5.0000	7470A	PPB	08 Feb 2018 09:21:48	S	Std
7.5 - 1	S7.5	27017	-	7.5000	7470A	PPB	08 Feb 2018 09:23:56	S	Std
10.0 - 1	S10	36625	-	10.0000	7470A	PPB	08 Feb 2018 09:30:58	S	Std
ICV90 - 1	ICV90	13365	3.6566	-	7470A	PPB	08 Feb 2018 09:33:28	U	SMPL
ICB90 - 1	ICB90	-81	-0.0334	-	7470A	PPB	08 Feb 2018 09:35:35	U	SMPL
CCV96 - 1	CCV96	18904	5.1767	-	7470A	PPB	08 Feb 2018 09:37:42	U	SMPL
CCB96 - 1	CCB96	-116	-0.0430	-	7470A	PPB	08 Feb 2018 09:39:48	U	SMPL
CRA - 1	CRA	770	0.2001	-	7470A	PPB	08 Feb 2018 09:41:56	U	SMPL
HighStd - 1	HighStd	37228	10.2053	-	7470A	PPB	08 Feb 2018 09:44:02	U	SMPL
ChkStd - 1	ChkStd	25335	6.9415	-	7470A	PPB	08 Feb 2018 09:46:09	U	SMPL
PB106380BL - 1	PBW	-6	-0.0128	-	7470A	PPB	08 Feb 2018 09:53:27	U	SMPL
PB106380BS - 1	LCSW	13490	3.6909	-	7470A	PPB	08 Feb 2018 09:55:34	U	SMPL
PB106380TB - 1	PB106341TB	-118	-0.0436	-	7470A	PPB	08 Feb 2018 09:57:42	U	SMPL
J1400-02 - 1	EO-01-020618-A	-8	-0.0134	-	7470A	PPB	08 Feb 2018 09:59:50	U	SMPL
J1400-04 - 1	EO-01-020618-B	-22	-0.0172	-	7470A	PPB	08 Feb 2018 10:01:58	U	SMPL
J1400-06 - 1	EO-01-020618-C	0	-0.0112	-	7470A	PPB	08 Feb 2018 10:04:07	U	SMPL
J1403-02 - 1	SU-04-020618	-25	-0.0181	-	7470A	PPB	08 Feb 2018 10:06:14	U	SMPL
CCV97 - 1	CCV97	18809	5.1506	-	7470A	PPB	08 Feb 2018 10:08:20	U	SMPL
CCB97 - 1	CCB97	-98	-0.0381	-	7470A	PPB	08 Feb 2018 10:10:26	U	SMPL
J1403-02DUP - 1	SU-04-020618DUP	-11	-0.0142	-	7470A	PPB	08 Feb 2018 10:12:34	U	SMPL
J1403-02MS - 1	SU-04-020618MS	12861	3.5183	-	7470A	PPB	08 Feb 2018 10:14:41	U	SMPL
J1403-02MSD - 1	SU-04-020618MSD	12736	3.4840	-	7470A	PPB	08 Feb 2018 10:16:47	U	SMPL
J1461-01 - 1	CHASSIS-WASH	-117	-0.0433	-	7470A	PPB	08 Feb 2018 10:18:54	U	SMPL
J1461-02 - 1	BUS-WASH	-56	-0.0266	-	7470A	PPB	08 Feb 2018 10:21:02	U	SMPL
J1461-03 - 1	OIL-WATER-SEPERATOR	-19	-0.0164	-	7470A	PPB	08 Feb 2018 10:23:10	U	SMPL
J1461-04 - 1	RECLAIMER-SLUDGE-PIT	-41	-0.0224	-	7470A	PPB	08 Feb 2018 10:25:18	U	SMPL
J1465-03 - 1	RT-3451	14	-0.0073	-	7470A	PPB	08 Feb 2018 10:27:26	U	SMPL
J1466-02 - 1	FRAC-TANK	-2	-0.0117	-	7470A	PPB	08 Feb 2018 10:29:35	U	SMPL
CCV98 - 1	CCV98	18861	5.1649	-	7470A	PPB	08 Feb 2018 10:31:42	U	SMPL
CCB98 - 1	CCB98	-100	-0.0386	-	7470A	PPB	08 Feb 2018 10:33:48	U	SMPL
J1403-02LX5 - 1		-42	-0.0227	-	7470A	PPB	08 Feb 2018 10:35:57	U	SMPL
J1403-02A - 1		13474	3.6865	-	7470A	PPB	08 Feb 2018 10:38:04	U	SMPL
CCV99 - 1	CCV99	18582	5.0883	-	7470A	PPB	08 Feb 2018 10:40:11	U	SMPL
CCB99 - 1	CCB99	-133	-0.0477	-	7470A	PPB	08 Feb 2018 10:42:18	U	SMPL