



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB93158

Temp.: 22.1 °C

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	02/07/2018	16:30
2	CAL2	1	Water			7.01	02/07/2018	16:31
3	CAL3	1	Water			10.02	02/07/2018	16:32
4	ICV	1	Water			7.00	02/07/2018	16:35
5	CCV1	1	Water			2.02	02/07/2018	16:45
6	J1482-01	1	Water			8.25	02/07/2018	16:46
7	J1490-06	1	Water			7.77	02/07/2018	16:47
8	J1490-06DUP	1	Water			7.78	02/07/2018	16:48
9	CCV2	1	Water			12.02	02/07/2018	16:49

LB 93158

WORKLIST(Hardcopy Internal Chain)

WorkList Name : phj1490 WorkList ID : 108340 Date : 2/7/2018 4:38:35 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/06/2018	Water	J1482-01-4	pH	Cool 4 deg C	LAND03	O53	GL-W-EFF-2-6-2018-01	02/06/2018	9040C
02/07/2018	Water	J1490-06-13	pH	Cool 4 deg C	PSEG02	O62	265578-SEGMENT-2C	02/07/2018	9040C

Date/Time 02-07-18 4:40 PM
 Received by: JP
 Relinquished by: CP

Date/Time 02-07-18 6:00 PM
 Received by: CP
 Relinquished by: JP