


7471B

LB 93162
INSTRUMENT ID: CV4

Linear

 μ Abs.:

42745



A= 0.0000e+000

B= 2.3533e-004

C= -1.6640e-002

Rho= 0.9999236

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.002	0.002	78	0.000	78				
0.2	0.200	0.193	-0.007	889	0.0 %	889				
2.5	2.500	2.567	0.067	10980	0.0 %	10980				
5.0	5.000	4.929	-0.071	21018	0.0 %	21018				
7.5	7.500	7.467	-0.033	31799	0.0 %	31799				
10	10.000	10.042	0.042	42745	0.0 %	42745				

LB93162 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 08 Feb 2018 13:23:15

Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	78	-	0.00007471B	-	PPB	08 Feb 2018 13:25:04	S	Std
0.2 - 1	50.2	889	-	0.20007471B	-	PPB	08 Feb 2018 13:27:21	S	Std
2.5 - 1	52.5	10980	-	2.50007471B	-	PPB	08 Feb 2018 13:29:39	S	Std
5.0 - 1	55	21018	-	5.00007471B	-	PPB	08 Feb 2018 13:31:56	S	Std
7.5 - 1	57.5	31799	-	7.50007471B	-	PPB	08 Feb 2018 13:34:13	S	Std
10 - 1	510	42745	-	10.00007471B	-	PPB	08 Feb 2018 13:39:13	S	Std
ICV91 - 1	ICV91	15714	3.6813	-7471B	-	PPB	08 Feb 2018 13:41:54	U	SMPL
ICB91 - 1	ICB91	-73	-0.0338	-7471B	-	PPB	08 Feb 2018 13:44:10	U	SMPL
CCV01 - 1	CCV01	21530	5.0500	-7471B	-	PPB	08 Feb 2018 13:46:26	U	SMPL
CCB01 - 1	CCB01	-111	-0.0428	-7471B	-	PPB	08 Feb 2018 13:48:41	U	SMPL
CRA - 1	CRA	876	0.1895	-7471B	-	PPB	08 Feb 2018 13:50:58	U	SMPL
HighStd - 1	HighStd	43104	10.1269	-7471B	-	PPB	08 Feb 2018 13:53:13	U	SMPL
ChkStd - 1	ChkStd	30392	7.1354	-7471B	-	PPB	08 Feb 2018 13:55:29	U	SMPL
PB106378BL - 1	PBS	-81	-0.0357	-7471B	-	PPB	08 Feb 2018 13:57:47	U	SMPL
PB106378BS - 1	LCSS	14256	3.3382	-7471B	-	PPB	08 Feb 2018 14:00:04	U	SMPL
J1469-01 - 1	WABUT-1	1580	0.3552	-7471B	-	PPB	08 Feb 2018 14:02:21	U	SMPL
J1469-03 - 1	WABUT-2	557	0.1144	-7471B	-	PPB	08 Feb 2018 14:07:15	U	SMPL
J1469-03DUP - 1	WABUT-2DUP	539	0.1102	-7471B	-	PPB	08 Feb 2018 14:09:32	U	SMPL
J1469-03MS - 1	WABUT-2MS	16626	3.8959	-7471B	-	PPB	08 Feb 2018 14:11:50	U	SMPL
J1469-03MSD - 1	WABUT-2MSD	17082	4.0032	-7471B	-	PPB	08 Feb 2018 14:14:06	U	SMPL
CCV02 - 1	CCV02	21127	4.9551	-7471B	-	PPB	08 Feb 2018 14:16:23	U	SMPL
CCB02 - 1	CCB02	-54	-0.0293	-7471B	-	PPB	08 Feb 2018 14:18:39	U	SMPL
J1469-05 - 1	WABUT-3	1232	0.2733	-7471B	-	PPB	08 Feb 2018 14:20:55	U	SMPL
J1469-07 - 1	WABUT-4	761	0.1624	-7471B	-	PPB	08 Feb 2018 14:23:11	U	SMPL
J1469-09 - 1	WABUT-5	1758	0.3971	-7471B	-	PPB	08 Feb 2018 14:25:27	U	SMPL
J1469-11 - 1	WABUT-6	2120	0.4823	-7471B	-	PPB	08 Feb 2018 14:27:44	U	SMPL
J1469-13 - 1	3N-1	1404	0.3138	-7471B	-	PPB	08 Feb 2018 14:30:00	U	SMPL
J1469-15 - 1	3N-2	1339	0.2985	-7471B	-	PPB	08 Feb 2018 14:32:18	U	SMPL
J1469-17 - 1	3N-3	1505	0.3375	-7471B	-	PPB	08 Feb 2018 14:34:35	U	SMPL
J1469-19 - 1	3N-4	980	0.2140	-7471B	-	PPB	08 Feb 2018 14:36:53	U	SMPL
J1470-01 - 1	3N-5	1350	0.3011	-7471B	-	PPB	08 Feb 2018 14:39:10	U	SMPL
J1470-03 - 1	3N-6	3833	0.8854	-7471B	-	PPB	08 Feb 2018 14:41:27	U	SMPL
CCV03 - 1	CCV03	21068	4.9412	-7471B	-	PPB	08 Feb 2018 14:43:43	U	SMPL
CCB03 - 1	CCB03	-55	-0.0296	-7471B	-	PPB	08 Feb 2018 14:45:58	U	SMPL
J1470-05 - 1	3N-7	1170	0.2587	-7471B	-	PPB	08 Feb 2018 14:48:15	U	SMPL
J1470-07 - 1	3N-8	1237	0.2745	-7471B	-	PPB	08 Feb 2018 14:50:31	U	SMPL
J1470-09 - 1	3N-9	2402	0.5486	-7471B	-	PPB	08 Feb 2018 14:52:47	U	SMPL
J1470-11 - 1	3N-10	1401	0.3131	-7471B	-	PPB	08 Feb 2018 14:55:04	U	SMPL
J1470-13 - 1	3N-11	1369	0.3055	-7471B	-	PPB	08 Feb 2018 14:57:20	U	SMPL
J1470-15 - 1	3N-12	1547	0.3474	-7471B	-	PPB	08 Feb 2018 14:59:38	U	SMPL
J1470-17 - 1	3N-13	1408	0.3147	-7471B	-	PPB	08 Feb 2018 15:01:54	U	SMPL
J1470-19 - 1	3N-14	1521	0.3413	-7471B	-	PPB	08 Feb 2018 15:04:12	U	SMPL
PB106379BL - 1	PBS	-11	-0.0192	-7471B	-	PPB	08 Feb 2018 15:06:30	U	SMPL
CCV04 - 1	CCV04	20749	4.8662	-7471B	-	PPB	08 Feb 2018 15:11:02	U	SMPL
CCB04 - 1	CCB04	-90	-0.0378	-7471B	-	PPB	08 Feb 2018 15:13:19	U	SMPL
J1404-01 - 1	TR-05-020718	1848	0.4182	-7471B	-	PPB	08 Feb 2018 15:15:35	U	SMPL
J1471-01 - 1	3N-15	1639	0.3691	-7471B	-	PPB	08 Feb 2018 15:17:51	U	SMPL
J1471-03 - 1	3N-16	1687	0.3804	-7471B	-	PPB	08 Feb 2018 15:20:08	U	SMPL
J1471-05 - 1	3N-17	2006	0.4554	-7471B	-	PPB	08 Feb 2018 15:22:24	U	SMPL
J1471-07 - 1	3N-18	2314	0.5279	-7471B	-	PPB	08 Feb 2018 15:24:41	U	SMPL
J1473-01 - 1	B1	1674	0.3773	-7471B	-	PPB	08 Feb 2018 15:26:58	U	SMPL
J1473-05 - 1	B2	828	0.1782	-7471B	-	PPB	08 Feb 2018 15:29:15	U	SMPL
J1473-09 - 1	B3	3639	0.8397	-7471B	-	PPB	08 Feb 2018 15:31:32	U	SMPL
J1473-13 - 1	B4	936	0.2036	-7471B	-	PPB	08 Feb 2018 15:33:50	U	SMPL
J1473-17 - 1	B5	609	0.1267	-7471B	-	PPB	08 Feb 2018 15:36:06	U	SMPL
CCV05 - 1	CCV05	21117	4.9528	-7471B	-	PPB	08 Feb 2018 15:38:22	U	SMPL
CCB05 - 1	CCB05	-75	-0.0343	-7471B	-	PPB	08 Feb 2018 15:40:39	U	SMPL
J1473-21 - 1	B6	694	0.1467	-7471B	-	PPB	08 Feb 2018 15:42:55	U	SMPL
J1473-25 - 1	B7	755	0.1610	-7471B	-	PPB	08 Feb 2018 15:45:12	U	SMPL
J1473-29 - 1	B8	661	0.1389	-7471B	-	PPB	08 Feb 2018 15:47:28	U	SMPL
J1473-33 - 1	B9	617	0.1286	-7471B	-	PPB	08 Feb 2018 15:49:44	U	SMPL
J1473-37 - 1	B10	997	0.2180	-7471B	-	PPB	08 Feb 2018 15:52:01	U	SMPL
J1473-41 - 1	B11	2246	0.5119	-7471B	-	PPB	08 Feb 2018 15:54:18	U	SMPL
J1473-41DUP - 1	B11DUP	1994	0.4526	-7471B	-	PPB	08 Feb 2018 15:56:35	U	SMPL
J1473-41MS - 1	B11MS	18500	4.3369	-7471B	-	PPB	08 Feb 2018 15:58:52	U	SMPL
J1473-41MSD - 1	B11MSD	18447	4.3244	-7471B	-	PPB	08 Feb 2018 16:04:03	U	SMPL
J1485-01 - 1	RO-201	20348	4.7718	-7471B	-	PPB	08 Feb 2018 16:06:19	U	SMPL
CCV06 - 1	CCV06	20795	4.8770	-7471B	-	PPB	08 Feb 2018 16:09:51	U	SMPL
CCB06 - 1	CCB06	-82	-0.0359	-7471B	-	PPB	08 Feb 2018 16:12:07	U	SMPL
J1485-03 - 1	RO-215	16812	3.9397	-7471B	-	PPB	08 Feb 2018 16:14:24	U	SMPL
PB106383BL - 1	PBS	-61	-0.0310	-7471B	-	PPB	08 Feb 2018 16:16:41	U	SMPL
PB106383BS - 1	LCSS	14857	3.4796	-7471B	-	PPB	08 Feb 2018 16:21:40	U	SMPL
J1475-11 - 1	COMP-1	1895	0.4293	-7471B	-	PPB	08 Feb 2018 16:23:56	U	SMPL
J1475-12 - 1	COMP-2	8138	1.8985	-7471B	-	PPB	08 Feb 2018 16:26:12	U	SMPL
J1475-13 - 1	COMP-3	1925	0.4364	-7471B	-	PPB	08 Feb 2018 16:28:29	U	SMPL
J1476-01 - 1	GL-WC-01	243438	57.2711	-7471B	-	PPB	08 Feb 2018 16:39:35	U	SMPL
J1476-01DUP - 1	GL-WC-01DUP	241022	56.7025	-7471B	-	PPB	08 Feb 2018 16:41:52	U	SMPL
J1476-01MS - 1	GL-WC-01MS	255160	60.0296	-7471B	-	PPB	08 Feb 2018 16:44:27	U	SMPL
J1476-01MSD - 1	GL-WC-01MSD	255675	60.1508	-7471B	-	PPB	08 Feb 2018 16:52:00	U	SMPL
CCV07 - 1	CCV07	20079	4.7085	-7471B	-	PPB	08 Feb 2018 16:54:16	U	SMPL
CCB07 - 1	CCB07	-196	-0.0628	-7471B	-	PPB	08 Feb 2018 16:56:53	U	SMPL
J1476-03 - 1	GL-WC-02	12785	2.9920	-7471B	-	PPB	08 Feb 2018 16:59:12	U	SMPL
J1476-05 - 1	GL-WC-03	132720	31.2160	-7471B	-	PPB	08 Feb 2018 17:01:29	U	SMPL
J1490-03 - 1	2ND-ST-CONCRETE	10045	2.3472	-7471B	-	PPB	08 Feb 2018 17:03:46	U	SMPL
J1491-01 - 1	1-1-4-DENSE-GRADE-AGG-RCA	4298	0.9948	-7471B	-	PPB	08 Feb 2018 17:06:07	U	SMPL
J1492-01 - 1	1-1-4-DENSE-GRADE-AGG-RCA	2132	0.4851	-7471B	-	PPB	08 Feb 2018 17:08:24	U	SMPL
J1493-01 - 1	1-1-4-DENSE-GRADE-AGG-RCA	2964	0.6809	-7471B	-	PPB	08 Feb 2018 17:10:40	U	SMPL
PB106379BS - 1	LCSS	15100	3.5368	-7471B	-	PPB	08 Feb 2018 17:12:58	U	SMPL
CCV08 - 1	CCV08	20794	4.8768	-7471B	-	PPB	08 Feb 2018 17:15:15	U	SMPL
CCB08 - 1	CCB08	-125	-0.0461	-7471B	-	PPB	08 Feb 2018 17:17:33	U	SMPL
J1469-03LX5 - 1		110	0.0092	-7471B	-	PPB	08 Feb 2018 17:19:50	U	SMPL
J1469-03A - 1		15808	3.7034	-7471B	-	PPB	08 Feb 2018 17:22:07	U	SMPL
J1473-41LX5 - 1		326	0.0601	-7471B	-	PPB	08 Feb 2018 17:24:24	U	SMPL

LB93162 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 08 Feb 2018 13:23:15

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
J1473-41A - 1		17203	4.0317		-7471B	PPB	08 Feb 2018 17:26:41	U	SMPL
J1476-01LX5 - 1		48528	11.4033		-7471B	PPB	08 Feb 2018 17:28:58	U	SMPL
J1476-01A - 1		251703	59.2160		-7471B	PPB	08 Feb 2018 17:31:15	U	SMPL
CCV09 - 1	CCV09	20408	4.7859		-7471B	PPB	08 Feb 2018 17:33:33	U	SMPL
CCB09 - 1	CCB09	-227	-0.0701		-7471B	PPB	08 Feb 2018 17:36:10	U	SMPL
J1476-01DLX10 - 1		24170	5.6712		-7471B	PPB	08 Feb 2018 17:38:28	U	SMPL
J1476-01DUPDLX10 - 1		24182	5.6741		-7471B	PPB	08 Feb 2018 17:40:46	U	SMPL
J1476-01MSDLX10 - 1		25243	5.9237		-7471B	PPB	08 Feb 2018 17:43:04	U	SMPL
J1476-01MSDDLX10 - 1		26603	6.2438		-7471B	PPB	08 Feb 2018 17:45:23	U	SMPL
J1476-01LDLX50 - 1		4318	0.9995		-7471B	PPB	08 Feb 2018 17:47:40	U	SMPL
J1476-01ADLX10 - 1		26283	6.1685		-7471B	PPB	08 Feb 2018 18:01:09	U	SMPL
J1476-05DLX5 - 1		23636	5.5456		-7471B	PPB	08 Feb 2018 18:03:25	U	SMPL
CCV10 - 1	CCV10	20836	4.8866		-7471B	PPB	08 Feb 2018 18:05:43	U	SMPL
CCB10 - 1	CCB10	-148	-0.0515		-7471B	PPB	08 Feb 2018 18:08:02	U	SMPL