



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB93171

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP61448
COD calibration std. 100 ppm	WP61444
COD calibration std. 10 ppm	WP61443
COD calibration std. 150 ppm	WP61445
COD calibration std. 50 ppm	WP61446
COD calibration std. 0 ppm	WP61442
COD Digestion Vials Low Level 0-150Mg/L	W2291
COD ICV-LCS std, 50ppm	WP61732
COD CCV std, 50ppm	WP61733

Temp In (C): <u>150</u>	Date In: <u>02/08/2018</u>	Time In: <u>09:10</u>
Temp Out (C): <u>150</u>	Date Out: <u>02/08/2018</u>	Time Out: <u>11:11</u>

Intercept: 0.1033

Slope: 1.0048

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	01/30/2018	12:20
2	CAL2	10	1	Water	10.000	01/30/2018	12:20
3	CAL3	50	1	Water	51.000	01/30/2018	12:21
4	CAL4	100	1	Water	100.000	01/30/2018	12:21
5	CAL5	150	1	Water	151.000	01/30/2018	12:22



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB93171

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.658	01/30/2018	12:22
2	ICB				1	Water	0.000	-0.103	01/30/2018	12:23
3	CCV1	50			1	Water	51.000	50.654	02/08/2018	12:10
4	CCB1				1	Water	0.000	-0.103	02/08/2018	12:10
5	LB93171BL				1	Water	0.000	-0.103	02/08/2018	12:11
6	LB93171BS	50			1	Water	50.000	49.658	02/08/2018	12:11
7	J1466-01				1	Water	112.000	111.362	02/08/2018	12:12
8	J1466-01DUP				1	Water	111.000	110.367	02/08/2018	12:12
9	J1466-01MS	50			1	Water	156.000	155.152	02/08/2018	12:13
10	J1466-01MSD	50			1	Water	157.000	156.147	02/08/2018	12:13
11	CCV2	50			1	Water	51.000	50.654	02/08/2018	12:14
12	CCB2				1	Water	0.000	-0.103	02/08/2018	12:14

2893271

WORKLIST(Hardcopy Internal Chain)

WorkList Name : COD020718 WorkList ID : 108326 Date : 2/7/2018 10:22:33 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/09/2018	Water	J1466-01	COD	Conc H2SO4 tr	PSEG01	J63	FRAC-TANK	02/06/2018	SM5220 D

Date/Time 02/08/18 9:00
Received by: IB
Relinquished by: JR

Date/Time 02/08/18 9:40
Received by: JR
Relinquished by: IB