

=====

Test results

Aquakem 7.2AQ1

1393176
Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

2/8/2018 16:18

Test: Total CN ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.949	0.0	0.091	
ICB1	0.499	0.0	0.004	
CCV1	240.887	0.0	0.223	
CCB1	0.478	0.0	0.004	
PB106411BL	0.276	0.0	0.004	
PB106411BS	96.492	0.0	0.092	
MIDPB106411	239.930	0.0	0.222	
J1161-10	1.887	0.0	0.005	
J1161-12	2.264	0.0	0.005	
CCV2	254.837	0.0	0.236	
CCB2	0.251	0.0	0.004	

N	11
Mean	84.886
SD	109.4523
CV%	128.94

AS

Aquakem v. 7.2AQ1

Results from time period:

Thu Feb 08 16:13:43 2018

Thu Feb 08 16:18:05 2018

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	0.1165	µg/l	2/8/2018 10:01:53	
5.OPPBCN	A	Total CN	P	4.941	µg/l	2/8/2018 10:01:54	
10PPBCN	A	Total CN	P	9.9492	µg/l	2/8/2018 10:01:55	
50PPBCN	A	Total CN	P	50.184	µg/l	2/8/2018 10:01:56	
100PPBCN	A	Total CN	P	100.3683	µg/l	2/8/2018 10:01:57	
250PPBCN	A	Total CN	P	249.0627	µg/l	2/8/2018 10:01:58	
500PPBCN	A	Total CN	P	500.3782	µg/l	2/8/2018 10:01:59	
ICV1	S	Total CN	P	95.9492	µg/l	2/8/2018 16:13:44	
ICB1	S	Total CN	P	0.4989	µg/l	2/8/2018 16:13:46	
CCV1	S	Total CN	P	240.8867	µg/l	2/8/2018 16:13:48	
CCB1	S	Total CN	P	0.478	µg/l	2/8/2018 16:13:49	
PB106411BL	S	Total CN	P	0.2765	µg/l	2/8/2018 16:13:51	
PB106411BS	S	Total CN	P	96.4921	µg/l	2/8/2018 16:13:52	
MIDPB106411	S	Total CN	P	239.9297	µg/l	2/8/2018 16:13:53	
J1161-10	S	Total CN	P	1.8872	µg/l	2/8/2018 16:18:02	
J1161-12	S	Total CN	P	2.264	µg/l	2/8/2018 16:18:03	
CCV2	S	Total CN	P	254.8367	µg/l	2/8/2018 16:18:04	
CCB2	S	Total CN	P	0.2511	µg/l	2/8/2018 16:18:05	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS

2/8/2018 10:09


Test Total CN

Accepted 2/8/2018 10:09

Factor 1095
 Bias 0.003

Coeff. of det. 0.999994

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	0.1165	0.0000	
2	5.0PPBCN	0.008	4.9410	5.0000	
3	10PPBCN	0.013	9.9492	10.0000	
4	50PPBCN	0.049	50.1840	50.0000	
5	100PPBCN	0.095	100.3683	100.0000	
6	250PPBCN	0.231	249.0627	250.0000	
7	500PPBCN	0.460	500.3782	500.0000	