



## Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB93178

Ambient Barometric Pressure (mmHg): 760

| Reagent/Standard | Lot/Log # |
|------------------|-----------|
| p-xylene (ICV)   | W2168     |

| Seq | LabID       | True Value °F | Dilution | Celsius °C | Result °F | Anal Date  | Anal Time |
|-----|-------------|---------------|----------|------------|-----------|------------|-----------|
| 1   | ICV         | 81            | 1        | 27.0       | 80.6      | 02/08/2018 | 11:02     |
| 2   | J1490-07    |               | 1        | >100.0     | >212.0    | 02/08/2018 | 11:32     |
| 3   | J1490-07DUP |               | 1        | >100.0     | >212.0    | 02/08/2018 | 11:55     |

Result = (Celsius \* 1.8) + 32

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP2818

WorkList ID : 108353

Date : 2/8/2018 7:47:40 AM

| Due Date   | Matrix | Sample   | Test | Flash Point | Preservative | Customer | Storage Location | Customer Sample   | Collect Date | Method |
|------------|--------|----------|------|-------------|--------------|----------|------------------|-------------------|--------------|--------|
| 02/11/2018 | Water  | J1490-07 | C    |             | Cool 4 deg C | PSEG02   | O62              | 265578-SEGMENT-2C | 02/07/2018   | 1010A  |

Date/Time 2/8/18 10:30 AM  
 Received by: KS  
 Relinquished by: CP

Date/Time 2/8/18 12:00 PM  
 Received by: CP  
 Relinquished by: KS