


7470A

LB 93183
INSTRUMENT ID: CV1

Linear

 μ Abs.:

37147



A= 0.0000e+000

B= 2.7024e-004

C= 4.5458e-002

Rho= 0.9998925

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.036	0.036	-36	0.000	-36				
0.2	0.200	0.225	0.025	665	0.0 %	665				
2.5	2.500	2.476	-0.024	8993	0.0 %	8993				
5.0	5.000	4.959	-0.041	18182	0.0 %	18182				
7.5	7.500	7.421	-0.079	27291	0.0 %	27291				
10.0	10.000	10.084	0.084	37147	0.0 %	37147				

LB93183 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 Feb 2018 09:20:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-36	-	0.00007470A	PPB		09 Feb 2018 09:58:09	S	Std
0.2 - 1	50.2	665	-	0.20007470A	PPB		09 Feb 2018 10:00:17	S	Std
2.5 - 1	52.5	8993	-	2.50007470A	PPB		09 Feb 2018 10:02:25	S	Std
5.0 - 1	55	18182	-	5.00007470A	PPB		09 Feb 2018 10:04:34	S	Std
7.5 - 1	57.5	27291	-	7.50007470A	PPB		09 Feb 2018 10:09:25	S	Std
10.0 - 1	511	37147	-	10.00007470A	PPB		09 Feb 2018 10:16:04	S	Std
ICV92 - 1	ICV92	14269	3.9015	-7470A	PPB		09 Feb 2018 10:18:57	U	SMPL
ICB92 - 1	ICB92	-163	0.0014	-7470A	PPB		09 Feb 2018 10:21:03	U	SMPL
CCV11 - 1	CCV11	19031	5.1883	-7470A	PPB		09 Feb 2018 10:23:11	U	SMPL
CCB11 - 1	CCB11	-157	0.0030	-7470A	PPB		09 Feb 2018 10:25:17	U	SMPL
CRA - 1	CRA	672	0.2271	-7470A	PPB		09 Feb 2018 10:27:25	U	SMPL
HighStd - 1	HighStd	37242	10.1096	-7470A	PPB		09 Feb 2018 10:29:31	U	SMPL
ChkStd - 1	ChkStd	25414	6.9133	-7470A	PPB		09 Feb 2018 10:31:38	U	SMPL
PB106414BL - 1	PBW	-247	-0.0213	-7470A	PPB		09 Feb 2018 10:33:48	U	SMPL
PB106414BS - 1	LCSW	14127	3.8631	-7470A	PPB		09 Feb 2018 10:35:56	U	SMPL
PB106414TB - 1	PB106372TB	-151	0.0047	-7470A	PPB		09 Feb 2018 10:38:04	U	SMPL
J1404-02 - 1	TR-05-020718	-85	0.0225	-7470A	PPB		09 Feb 2018 10:40:12	U	SMPL
J1473-04 - 1	B1	-97	0.0192	-7470A	PPB		09 Feb 2018 10:42:20	U	SMPL
J1473-04DUP - 1	B1DUP	-81	0.0236	-7470A	PPB		09 Feb 2018 10:44:28	U	SMPL
J1473-04MS - 1	B1MS	13407	3.6685	-7470A	PPB		09 Feb 2018 10:46:35	U	SMPL
CCV12 - 1	CCV12	18936	5.1627	-7470A	PPB		09 Feb 2018 10:48:42	U	SMPL
CCB12 - 1	CCB12	-151	0.0047	-7470A	PPB		09 Feb 2018 10:50:49	U	SMPL
J1473-04MSD - 1	B1MSD	13272	3.6321	-7470A	PPB		09 Feb 2018 10:52:56	U	SMPL
J1473-08 - 1	B2	-151	0.0047	-7470A	PPB		09 Feb 2018 10:55:03	U	SMPL
J1473-12 - 1	B3	-87	0.0219	-7470A	PPB		09 Feb 2018 10:57:10	U	SMPL
J1473-16 - 1	B4	-83	0.0230	-7470A	PPB		09 Feb 2018 10:59:17	U	SMPL
J1473-20 - 1	B5	-96	0.0195	-7470A	PPB		09 Feb 2018 11:01:24	U	SMPL
J1473-24 - 1	B6	-59	0.0295	-7470A	PPB		09 Feb 2018 11:03:32	U	SMPL
J1473-28 - 1	B7	-53	0.0311	-7470A	PPB		09 Feb 2018 11:05:40	U	SMPL
J1473-32 - 1	B8	-61	0.0290	-7470A	PPB		09 Feb 2018 11:07:49	U	SMPL
J1473-36 - 1	B9	-64	0.0282	-7470A	PPB		09 Feb 2018 11:09:57	U	SMPL
J1473-40 - 1	B10	-73	0.0257	-7470A	PPB		09 Feb 2018 11:12:04	U	SMPL
CCV13 - 1	CCV13	18709	5.1013	-7470A	PPB		09 Feb 2018 11:14:11	U	SMPL
CCB13 - 1	CCB13	-165	0.0009	-7470A	PPB		09 Feb 2018 11:16:18	U	SMPL
J1473-44 - 1	B11	-85	0.0225	-7470A	PPB		09 Feb 2018 11:18:26	U	SMPL
J1476-07 - 1	GL-WC-01	-79	0.0241	-7470A	PPB		09 Feb 2018 11:20:33	U	SMPL
J1476-08 - 1	GL-WC-02	-70	0.0265	-7470A	PPB		09 Feb 2018 11:22:40	U	SMPL
J1476-09 - 1	GL-WC-03	-46	0.0330	-7470A	PPB		09 Feb 2018 11:24:47	U	SMPL
J1485-02 - 1	RO-201	-41	0.0344	-7470A	PPB		09 Feb 2018 11:26:55	U	SMPL
J1485-04 - 1	RO-215	-75	0.0252	-7470A	PPB		09 Feb 2018 11:29:02	U	SMPL
J1487-01 - 1	DRUM-1-2	-50	0.0319	-7470A	PPB		09 Feb 2018 11:31:11	U	SMPL
J1490-04 - 1	2ND-ST-CONCRETE	-80	0.0238	-7470A	PPB		09 Feb 2018 11:33:18	U	SMPL
PB106415BL - 1	PBW	-40	0.0346	-7470A	PPB		09 Feb 2018 11:35:27	U	SMPL
PB106415BS - 1	LCSW	13247	3.6253	-7470A	PPB		09 Feb 2018 11:37:34	U	SMPL
CCV14 - 1	CCV14	18884	5.1486	-7470A	PPB		09 Feb 2018 11:39:41	U	SMPL
CCB14 - 1	CCB14	-165	0.0009	-7470A	PPB		09 Feb 2018 11:41:48	U	SMPL
PB106415TB - 1	PB106382TB	-63	0.0284	-7470A	PPB		09 Feb 2018 11:43:56	U	SMPL
J1490-07 - 1	265578-SEGMENT-2C	-41	0.0344	-7470A	PPB		09 Feb 2018 11:46:03	U	SMPL
J1490-07DUP - 1	265578-SEGMENT-2CDUP	-66	0.0276	-7470A	PPB		09 Feb 2018 11:48:10	U	SMPL
J1490-07MS - 1	265578-SEGMENT-2CMS	13183	3.6080	-7470A	PPB		09 Feb 2018 11:50:17	U	SMPL
J1490-07MSD - 1	265578-SEGMENT-2CMSD	13227	3.6199	-7470A	PPB		09 Feb 2018 11:52:25	U	SMPL
PB106416BL - 1	PBW	-183	-0.0040	-7470A	PPB		09 Feb 2018 11:54:32	U	SMPL
PB106416BS - 1	LCSW	13235	3.6221	-7470A	PPB		09 Feb 2018 11:56:40	U	SMPL
J1490-06 - 1	265578-SEGMENT-2C	-318	-0.0405	-7470A	PPB		09 Feb 2018 11:58:49	U	SMPL
J1490-06DUP - 1	265578-SEGMENT-2CDUP	-247	-0.0213	-7470A	PPB		09 Feb 2018 12:00:58	U	SMPL
J1490-06MS - 1	265578-SEGMENT-2CMS	13029	3.5664	-7470A	PPB		09 Feb 2018 12:03:05	U	SMPL
CCV15 - 1	CCV15	18072	4.9292	-7470A	PPB		09 Feb 2018 12:05:12	U	SMPL
CCB15 - 1	CCB15	-170	-0.0005	-7470A	PPB		09 Feb 2018 12:07:19	U	SMPL
J1490-06MSD - 1	265578-SEGMENT-2CMSD	13013	3.5621	-7470A	PPB		09 Feb 2018 12:09:27	U	SMPL
J1473-04LX5 - 1		-146	0.0060	-7470A	PPB		09 Feb 2018 12:11:34	U	SMPL
J1473-04A - 1		13435	3.6761	-7470A	PPB		09 Feb 2018 12:13:41	U	SMPL
J1490-07LX5 - 1		-148	0.0055	-7470A	PPB		09 Feb 2018 12:15:48	U	SMPL
J1490-07A - 1		13098	3.5850	-7470A	PPB		09 Feb 2018 12:17:55	U	SMPL
J1490-06LX5 - 1		-155	0.0036	-7470A	PPB		09 Feb 2018 12:20:02	U	SMPL
J1490-06A - 1		12993	3.5567	-7470A	PPB		09 Feb 2018 12:22:11	U	SMPL
CCV16 - 1	CCV16	17965	4.9003	-7470A	PPB		09 Feb 2018 12:24:19	U	SMPL
CCB16 - 1	CCB16	-196	-0.0075	-7470A	PPB		09 Feb 2018 12:26:28	U	SMPL