



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB93194

BalanceID: WC SC-4

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP61448
COD calibration std. 100 ppm	WP61444
COD calibration std. 10 ppm	WP61443
COD calibration std. 150 ppm	WP61445
COD calibration std. 50 ppm	WP61446
COD calibration std. 0 ppm	WP61442
COD Digestion Vials Low Level 0-150Mg/L	W2291
COD ICV-LCS std, 50ppm	WP61787
COD CCV std, 50ppm	WP61788
COD LOQ std, 10.0PPM	WP61799
COD LOD std, 5ppm	WP61801

Temp In (C): <u>150</u>	Date In: <u>02/09/2018</u>	Time In: <u>09:00</u>
Temp Out (C): <u>150</u>	Date Out: <u>02/09/2018</u>	Time Out: <u>11:01</u>

Intercept: 0.1033

Slope: 1.0048

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	01/30/2018	12:20
2	CAL2	10	1	Water	10.000	01/30/2018	12:20
3	CAL3	50	1	Water	51.000	01/30/2018	12:21
4	CAL4	100	1	Water	100.000	01/30/2018	12:21
5	CAL5	150	1	Water	151.000	01/30/2018	12:22



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB93194

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.658	01/30/2018	12:22
2	ICB				1	Water	0.000	-0.103	01/30/2018	12:23
3	CCV1	50			1	Water	51.000	50.654	02/09/2018	12:03
4	CCB1				1	Water	0.000	-0.103	02/09/2018	12:03
5	LB93194BL		1.00	50	1	Solid	0.000	-0.103	02/09/2018	12:04
6	LB93194BS	50	1.01	50	1	Solid	50.000	49.658	02/09/2018	12:04
7	J1161-01		1.00	50	1	Solid	7.000	6.864	02/09/2018	12:05
8	J1161-02		1.01	50	1	Solid	12.000	11.840	02/09/2018	12:05
9	J1161-03		1.00	50	1	Solid	5.000	4.873	02/09/2018	12:06
10	CCV2	50			1	Water	51.000	50.654	02/09/2018	12:06
11	CCB2				1	Water	0.000	-0.103	02/09/2018	12:07