


7471B

LB 93195
INSTRUMENT ID: CV2

Linear

 μ Abs.:

42310



A= 0.0000e+000

B= 2.3749e-004

C= -2.2036e-002

Rho= 0.9996537

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.033	-0.033	-47	0.000	-47				
0.2	0.200	0.179	-0.021	848	0.0 %	848				
2.5	2.500	2.521	0.021	10706	0.0 %	10706				
5.0	5.000	5.168	0.168	21852	0.0 %	21852				
7.5	7.500	7.339	-0.161	30996	0.0 %	30996				
10	10.000	10.026	0.026	42310	0.0 %	42310				

LB93195 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 09 Feb 2018 16:11:27

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-47	-	0.0000	7471B	PPB	09 Feb 2018 16:13:21	S	Std
0.2 - 1	50.2	848	-	0.2000	7471B	PPB	09 Feb 2018 16:15:39	S	Std
2.5 - 1	52.5	10706	-	2.5000	7471B	PPB	09 Feb 2018 16:17:56	S	Std
5.0 - 1	55	21852	-	5.0000	7471B	PPB	09 Feb 2018 16:20:14	S	Std
7.5 - 1	57.5	30996	-	7.5000	7471B	PPB	09 Feb 2018 16:22:31	S	Std
10 - 1	511	42310	-	10.0000	7471B	PPB	09 Feb 2018 16:27:25	S	Std
ICV94 - 1	ICV94	16656	3.9337	-7471B	PPB	PPB	09 Feb 2018 16:30:02	U	SMPL
ICB94 - 1	ICB94	-195	-0.0683	-7471B	PPB	PPB	09 Feb 2018 16:32:18	U	SMPL
CCV20 - 1	CCV20	21803	5.1560	-7471B	PPB	PPB	09 Feb 2018 16:34:34	U	SMPL
CCB20 - 1	CCB20	-202	-0.0700	-7471B	PPB	PPB	09 Feb 2018 16:36:50	U	SMPL
CRA - 1	CRA	795	0.1668	-7471B	PPB	PPB	09 Feb 2018 16:39:06	U	SMPL
HighStd - 1	HighStd	41822	9.9104	-7471B	PPB	PPB	09 Feb 2018 16:41:21	U	SMPL
ChkStd - 1	ChkStd	29464	6.9755	-7471B	PPB	PPB	09 Feb 2018 16:43:38	U	SMPL
PB106421BL - 1	PBS	-252	-0.0819	-7471B	PPB	PPB	09 Feb 2018 16:45:56	U	SMPL
PB106421BS - 1	LCSS	15518	3.6634	-7471B	PPB	PPB	09 Feb 2018 16:48:14	U	SMPL
J1499-01 - 1	RO-154	1134	0.2473	-7471B	PPB	PPB	09 Feb 2018 16:50:31	U	SMPL
J1499-01DUP - 1	RO-154DUP	1053	0.2280	-7471B	PPB	PPB	09 Feb 2018 16:52:48	U	SMPL
J1499-01MS - 1	RO-154MS	16261	3.8398	-7471B	PPB	PPB	09 Feb 2018 16:55:05	U	SMPL
J1499-01MSD - 1	RO-154MSD	16269	3.8417	-7471B	PPB	PPB	09 Feb 2018 16:57:23	U	SMPL
J1501-01 - 1	CO-6817	38	-0.0130	-7471B	PPB	PPB	09 Feb 2018 16:59:39	U	SMPL
CCV21 - 1	CCV21	21852	5.1677	-7471B	PPB	PPB	09 Feb 2018 17:01:55	U	SMPL
CCB21 - 1	CCB21	-162	-0.0605	-7471B	PPB	PPB	09 Feb 2018 17:04:12	U	SMPL
J1506-01 - 1	B-20(2-4)	2735	0.6275	-7471B	PPB	PPB	09 Feb 2018 17:06:28	U	SMPL
J1506-02 - 1	B-20(8-10)	8082	1.8974	-7471B	PPB	PPB	09 Feb 2018 17:08:44	U	SMPL
J1506-03 - 1	B-20(14-16)	200	0.0255	-7471B	PPB	PPB	09 Feb 2018 17:11:01	U	SMPL
J1506-04 - 1	B-21(2-4)	15665	3.6983	-7471B	PPB	PPB	09 Feb 2018 17:13:17	U	SMPL
J1506-05 - 1	B-21(8-10)	963	0.2067	-7471B	PPB	PPB	09 Feb 2018 17:15:34	U	SMPL
J1506-06 - 1	B-21(14-16)	7954	1.8670	-7471B	PPB	PPB	09 Feb 2018 17:17:51	U	SMPL
J1506-07 - 1	B-22(2-4)	2926	0.6729	-7471B	PPB	PPB	09 Feb 2018 17:20:08	U	SMPL
J1506-08 - 1	B-22(8-10)	3873	0.8978	-7471B	PPB	PPB	09 Feb 2018 17:22:25	U	SMPL
J1506-09 - 1	B-22(14-16)	384	0.0692	-7471B	PPB	PPB	09 Feb 2018 17:24:43	U	SMPL
J1507-01 - 1	VNJ-214	3389	0.7828	-7471B	PPB	PPB	09 Feb 2018 17:26:59	U	SMPL
CCV22 - 1	CCV22	21979	5.1978	-7471B	PPB	PPB	09 Feb 2018 17:29:16	U	SMPL
CCB22 - 1	CCB22	-168	-0.0619	-7471B	PPB	PPB	09 Feb 2018 17:31:32	U	SMPL
J1510-03 - 1	MADISON-MH	155	0.0148	-7471B	PPB	PPB	09 Feb 2018 17:33:49	U	SMPL
J1511-01 - 1	MM-1	47240	11.1971	-7471B	PPB	PPB	09 Feb 2018 17:36:05	U	SMPL
J1511-02 - 1	MM-2	27408	6.4872	-7471B	PPB	PPB	09 Feb 2018 17:38:21	U	SMPL
J1511-03 - 1	MM-3	88315	20.9522	-7471B	PPB	PPB	09 Feb 2018 17:40:40	U	SMPL
J1512-01 - 1	RO-2334	3544	0.8196	-7471B	PPB	PPB	09 Feb 2018 17:42:59	U	SMPL
J1512-04 - 1	RO-3091	380	0.0682	-7471B	PPB	PPB	09 Feb 2018 17:45:18	U	SMPL
J1512-07 - 1	RO-4252	951	0.2038	-7471B	PPB	PPB	09 Feb 2018 17:47:36	U	SMPL
J1512-10 - 1	RO-4632	1127	0.2456	-7471B	PPB	PPB	09 Feb 2018 17:49:53	U	SMPL
J1499-01LX5 - 1		138	0.0107	-7471B	PPB	PPB	09 Feb 2018 17:52:11	U	SMPL
J1499-01A - 1		17259	4.0769	-7471B	PPB	PPB	09 Feb 2018 17:54:27	U	SMPL
CCV23 - 1	CCV23	21891	5.1769	-7471B	PPB	PPB	09 Feb 2018 17:56:44	U	SMPL
CCB23 - 1	CCB23	-161	-0.0603	-7471B	PPB	PPB	09 Feb 2018 17:59:00	U	SMPL
J1511-01DLX2 - 1		23625	5.5887	-7471B	PPB	PPB	09 Feb 2018 18:01:17	U	SMPL
J1511-03DLX5 - 1		17782	4.2011	-7471B	PPB	PPB	09 Feb 2018 18:03:34	U	SMPL
CCV24 - 1	CCV24	22006	5.2042	-7471B	PPB	PPB	09 Feb 2018 18:05:51	U	SMPL
CCB24 - 1	CCB24	-191	-0.0674	-7471B	PPB	PPB	09 Feb 2018 18:08:08	U	SMPL