



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB93203

Temp.: 23.4 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/09/2018	17:35
2	CAL2	1	Water			7.00	02/09/2018	17:36
3	CAL3	1	Water			10.02	02/09/2018	17:37
4	ICV	1	Water			7.01	02/09/2018	17:39
5	CCV1	1	Water			2.02	02/09/2018	17:40
6	J1516-04	1	Solid	20.02	20	6.39	02/09/2018	17:41
7	J1517-02	1	Solid	20.04	20	10.25	02/09/2018	17:42
8	J1517-04	1	Solid	20.03	20	10.66	02/09/2018	17:44
9	J1517-06	1	Solid	20.05	20	11.01	02/09/2018	17:46
10	J1517-14	1	Solid	20.02	20	7.66	02/09/2018	17:48
11	J1517-22	1	Solid	20.01	20	6.94	02/09/2018	17:50
12	J1520-03	1	Solid	20.03	20	8.00	02/09/2018	17:51
13	J1526-03	1	Solid	20.02	20	8.04	02/09/2018	18:00
14	J1526-03DUP	1	Solid	20.04	20	8.05	02/09/2018	18:01
15	CCV2	1	Water			12.01	02/09/2018	18:02

LB 93203

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY J1527 WorkList ID : 108436

Date : 2/9/2018 5:12:08 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/09/2018	Solid	J1516-04 -A	Corrosivity	Cool 4 deg C	PSEG01	O51	WAVE-1	02/09/2018	9045C
02/09/2018	Solid	J1517-02 -B	Corrosivity	Cool 4 deg C	PSEG01	O51	CO-E	02/09/2018	9045C
02/09/2018	Solid	J1517-04 -A	Corrosivity	Cool 4 deg C	PSEG01	O51	CO-M	02/09/2018	9045C
02/09/2018	Solid	J1517-06 -A	Corrosivity	Cool 4 deg C	PSEG01	O51	CO-W	02/09/2018	9045C
02/09/2018	Solid	J1517-14 -A	Corrosivity	Cool 4 deg C	PSEG01	O51	TP-1	02/09/2018	9045C
02/09/2018	Solid	J1517-22 -A	Corrosivity	Cool 4 deg C	PSEG01	O51	TP-2	02/09/2018	9045C
02/09/2018	Solid	J1520-03 -A	Corrosivity	Cool 4 deg C	PSEG02	P51	OILY-STONES	02/09/2018	9045C
02/09/2018	Solid	J1526-03 -B	Corrosivity	Cool 4 deg C	PSEG01	P52	RO-1-2	02/09/2018	9045C

Date/Time 02-09-18 5:20 PM
 Received by: JP
 Relinquished by: CP

Date/Time 02-09-18 6:40 PM
 Received by: CP
 Relinquished by: JP