



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB93205

Temp.: 23.4 °C

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/09/2018	17:35
2	CAL2	1	Water			7.00	02/09/2018	17:36
3	CAL3	1	Water			10.02	02/09/2018	17:37
4	ICV	1	Water			7.01	02/09/2018	17:39
5	CCV1	1	Water			2.01	02/09/2018	18:10
6	J1527-01	1	Water			6.25	02/09/2018	18:11
7	J1527-01DUP	1	Water			6.26	02/09/2018	18:12
8	CCV2	1	Water			12.02	02/09/2018	18:15

2393205

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH J1526

WorkList ID : 108438

Date : 2/9/2018 5:13:42 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/09/2018	Water	J1527-01 -A	pH	Cool 4 deg C	PSEG01	P52	OILY-WATER	02/09/2018	9040C

Date/Time 02-09-18 5:20 PM
Received by: JD
Relinquished by: JD

Date/Time 02-09-18 6:35 AM
Received by: CD
Relinquished by: JF