


7470A

LB93227
INSTRUMENT ID: CY1

Linear

 μ Abs.:

37263



A= 0.0000e+000

B= 2.7102e-004

C= 3.8248e-002

Rho= 0.9996870

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.056	0.056	66	0.000	66				
0.2	0.200	0.202	0.002	605	0.0 %	605				
2.5	2.500	2.514	0.014	9135	0.0 %	9135				
5.0	5.000	4.955	-0.045	18141	0.0 %	18141				
7.5	7.500	7.336	-0.164	26927	0.0 %	26927				
10.0	10.000	10.137	0.137	37263	0.0 %	37263				

LB93227 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Feb 2018 15:40:15

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	66	-	0.0000	7470A	PPB	12 Feb 2018 15:45:37	S	Std
0.2 - 1	50.2	605	-	0.2000	7470A	PPB	12 Feb 2018 15:47:45	S	Std
2.5 - 1	52.5	9135	-	2.5000	7470A	PPB	12 Feb 2018 15:49:53	S	Std
5.0 - 1	55	18141	-	5.0000	7470A	PPB	12 Feb 2018 15:52:02	S	Std
7.5 - 1	57.5	26927	-	7.5000	7470A	PPB	12 Feb 2018 15:54:09	S	Std
10.0 - 1	510	37263	-	10.0000	7470A	PPB	12 Feb 2018 16:01:08	S	Std
ICV95 - 1	ICV95	13523	3.7032	-	7470A	PPB	12 Feb 2018 16:05:34	U	SMPL
ICB95 - 1	ICB95	-53	0.0239	-	7470A	PPB	12 Feb 2018 16:07:40	U	SMPL
CCV25 - 1	CCV25	18002	4.9171	-	7470A	PPB	12 Feb 2018 16:09:47	U	SMPL
CCB25 - 1	CCB25	-68	0.0198	-	7470A	PPB	12 Feb 2018 16:11:53	U	SMPL
CRA - 1	CRA	628	0.2084	-	7470A	PPB	12 Feb 2018 16:14:01	U	SMPL
HighStd - 1	HighStd	37075	10.0861	-	7470A	PPB	12 Feb 2018 16:16:07	U	SMPL
ChkStd - 1	ChkStd	24611	6.7082	-	7470A	PPB	12 Feb 2018 16:18:14	U	SMPL
PB106447BL - 1	PBW	-169	-0.0076	-	7470A	PPB	12 Feb 2018 16:20:23	U	SMPL
PB106447BS - 1	LCSW	14033	3.8414	-	7470A	PPB	12 Feb 2018 16:22:31	U	SMPL
PB106447TB - 1	PB106419TB	-97	0.0120	-	7470A	PPB	12 Feb 2018 16:24:39	U	SMPL
J1499-03 - 1	RO-154	-60	0.0220	-	7470A	PPB	12 Feb 2018 16:26:46	U	SMPL
J1499-03DUP - 1	RO-154DUP	-49	0.0250	-	7470A	PPB	12 Feb 2018 16:28:55	U	SMPL
J1499-03MS - 1	RO-154MS	13624	3.7306	-	7470A	PPB	12 Feb 2018 16:31:03	U	SMPL
J1499-03MSD - 1	RO-154MSD	13160	3.6048	-	7470A	PPB	12 Feb 2018 16:33:10	U	SMPL
CCV26 - 1	CCV26	18352	5.0119	-	7470A	PPB	12 Feb 2018 16:35:17	U	SMPL
CCB26 - 1	CCB26	-71	0.0190	-	7470A	PPB	12 Feb 2018 16:37:24	U	SMPL
J1501-02 - 1	CO-6817	-30	0.0301	-	7470A	PPB	12 Feb 2018 16:39:31	U	SMPL
J1507-03 - 1	VNJ-214	-45	0.0261	-	7470A	PPB	12 Feb 2018 16:41:38	U	SMPL
J1510-06 - 1	MADISON-MH	-26	0.0312	-	7470A	PPB	12 Feb 2018 16:43:45	U	SMPL
J1511-04 - 1	MM-1	-12	0.0350	-	7470A	PPB	12 Feb 2018 16:45:52	U	SMPL
J1511-05 - 1	MM-2	54	0.0529	-	7470A	PPB	12 Feb 2018 16:47:59	U	SMPL
J1511-06 - 1	MM-3	58	0.0540	-	7470A	PPB	12 Feb 2018 16:50:07	U	SMPL
J1512-03 - 1	RO-2334	-17	0.0336	-	7470A	PPB	12 Feb 2018 16:52:15	U	SMPL
J1512-06 - 1	RO-3091	-9	0.0358	-	7470A	PPB	12 Feb 2018 16:54:23	U	SMPL
J1512-09 - 1	RO-4252	-10	0.0355	-	7470A	PPB	12 Feb 2018 16:56:32	U	SMPL
J1512-12 - 1	RO-4632	-18	0.0334	-	7470A	PPB	12 Feb 2018 16:58:39	U	SMPL
CCV27 - 1	CCV27	18237	4.9807	-	7470A	PPB	12 Feb 2018 17:00:46	U	SMPL
CCB27 - 1	CCB27	-64	0.0209	-	7470A	PPB	12 Feb 2018 17:02:53	U	SMPL
J1512-13 - 1	Oily-Debris	37	0.0483	-	7470A	PPB	12 Feb 2018 17:05:01	U	SMPL
J1499-03LX5 - 1		33	0.0472	-	7470A	PPB	12 Feb 2018 17:07:08	U	SMPL
J1499-03A - 1		13040	3.5723	-	7470A	PPB	12 Feb 2018 17:09:14	U	SMPL
CCV28 - 1	CCV28	18398	5.0244	-	7470A	PPB	12 Feb 2018 17:11:21	U	SMPL
CCB28 - 1	CCB28	-54	0.0236	-	7470A	PPB	12 Feb 2018 17:13:29	U	SMPL