

Daily Analysis Runlog For Sequence/QC Batch ID # LB93231

Review By		Review On				
SubDirectory		LB93231		Test : Cyanide		
STD. NAME		STD REF.#				
ICAL Standard		WP61806,WP61807,WP61808,WP61809,WP61810,WP61811,WP61815				
ICV Standard		W1542				
CCV Standard		WP61812				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		WP58910				
Chk Standard		WP61814,WP59958,WP60878				

Sr#	SampleID	ClientID	QcType	Date	Comment	Status
1	0.0PPBCN	0.0PPBCN	CAL1	02/12/18 08:50		OK
2	5.0PPBCN	5.0PPBCN	CAL2	02/12/18 08:50		OK
3	10PPBCN	10PPBCN	CAL3	02/12/18 08:50		OK
4	50PPBCN	50PPBCN	CAL4	02/12/18 08:50		OK
5	100PPBCN	100PPBCN	CAL5	02/12/18 08:50		OK
6	250PPBCN	250PPBCN	CAL6	02/12/18 08:50		OK
7	500PPBCN	500PPBCN	CAL7	02/12/18 08:50		OK
8	ICV1	ICV1	ICV	02/12/18 15:24		OK
9	ICB1	ICB1	ICB	02/12/18 15:24		OK
10	CCV1	CCV1	CCV	02/12/18 15:24		OK
11	CCB1	CCB1	CCB	02/12/18 15:24		OK
12	PB106455BL	PB106455BL	MB	02/12/18 15:44		OK
13	PB106455BS	PB106455BS	LCS	02/12/18 15:44		OK
14	LOWPB106455	LOWPB106455	SAM	02/12/18 15:44		OK
15	HIGHPB106455	HIGHPB106455	SAM	02/12/18 15:44		OK
16	J1516-01	WAVE-1	SAM	02/12/18 15:44		OK
17	J1516-01DUP	WAVE-1DUP	DUP	02/12/18 15:44		OK
18	J1516-01MS	WAVE-1MS	MS	02/12/18 15:44		OK
19	J1516-01MSD	WAVE-1MSD	MSD	02/12/18 15:44		OK
20	J1517-01	CO-E	SAM	02/12/18 15:44		OK
21	J1517-03	CO-M	SAM	02/12/18 15:44		OK

Daily Analysis Runlog For Sequence/QC Batch ID # LB93231

Review By		Review On				
SubDirectory		LB93231		Test : Cyanide		
STD. NAME		STD REF.#				
ICAL Standard		WP61806,WP61807,WP61808,WP61809,WP61810,WP61811,WP61815				
ICV Standard		W1542				
CCV Standard		WP61812				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		WP58910				
Chk Standard		WP61814,WP59958,WP60878				
22	CCV2	CCV2	CCV	02/12/18 15:44		OK
23	CCB2	CCB2	CCB	02/12/18 15:52		OK
24	J1517-05	CO-W	SAM	02/12/18 15:52		OK
25	J1517-07	TP-1	SAM	02/12/18 15:52		OK
26	J1517-15	TP-2	SAM	02/12/18 15:52		OK
27	J1520-01	OILY-STONES	SAM	02/12/18 15:52		OK
28	J1525-01	SAMPLE-1	SAM	02/12/18 15:52		OK
29	J1525-03	SAMPLE-2	SAM	02/12/18 15:52		OK
30	J1525-05	SAMPLE-3	SAM	02/12/18 15:52		OK
31	J1526-01	RO-1-2	SAM	02/12/18 15:52		OK
32	PB106470BL	PB106470BL	MB	02/12/18 15:52		OK
33	PB106470BS	PB106470BS	LCS	02/12/18 15:52		OK
34	CCV3	CCV3	CCV	02/12/18 15:58		OK
35	CCB3	CCB3	CCB	02/12/18 15:58		OK
36	J1527-01	OILY-WATER	SAM	02/12/18 15:58		OK
37	J1527-01DUP	OILY-WATERDUP	DUP	02/12/18 15:58		OK
38	J1527-01MS	OILY-WATERMS	MS	02/12/18 15:58		OK
39	J1527-01MSD	OILY-WATERMSD	MSD	02/12/18 15:58		OK
40	CCV4	CCV4	CCV	02/12/18 15:58		OK
41	CCB4	CCB4	CCB	02/12/18 15:58		OK



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB93231	Date Received:	
Client Sample ID:	HIGHPB106455	SDG No.:	LB93231
Lab Sample ID:	HIGHPB106455	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.5		1	0.003		0.005	mg/L		02/12/2018	9012B

Report of Analysis

Client:		Date Collected:	2/9/2018 12:00:00 AM
Project:	LB93231	Date Received:	2/9/2018 12:00:00 AM
Client Sample ID:	WAVE-1	SDG No.:	LB93231
Lab Sample ID:	J1516-01	Matrix:	Solid
Level (low/med):	low	% Solid:	81.1

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.0069	U	1	0.04		0.3	mg/Kg	02/12/2018	02/12/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	2/9/2018 12:00:00 AM
Project:	LB93231	Date Received:	2/9/2018 12:00:00 AM
Client Sample ID:	CO-E	SDG No.:	LB93231
Lab Sample ID:	J1517-01	Matrix:	Solid
Level (low/med):	low	% Solid:	100

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.097	J	1	0.033		0.25	mg/Kg	02/12/2018	02/12/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	2/9/2018 12:00:00 AM
Project:	LB93231	Date Received:	2/9/2018 12:00:00 AM
Client Sample ID:	CO-M	SDG No.:	LB93231
Lab Sample ID:	J1517-03	Matrix:	Solid
Level (low/med):	low	% Solid:	100

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.039	J	1	0.031		0.24	mg/Kg	02/12/2018	02/12/2018	9012B

Report of Analysis

Client:		Date Collected:	2/9/2018 12:00:00 AM
Project:	LB93231	Date Received:	2/9/2018 12:00:00 AM
Client Sample ID:	CO-W	SDG No.:	LB93231
Lab Sample ID:	J1517-05	Matrix:	Solid
Level (low/med):	low	% Solid:	100

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.074	J	1	0.032		0.24	mg/Kg	02/12/2018	02/12/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	2/9/2018 12:00:00 AM
Project:	LB93231	Date Received:	2/9/2018 12:00:00 AM
Client Sample ID:	TP-1	SDG No.:	LB93231
Lab Sample ID:	J1517-07	Matrix:	Solid
Level (low/med):	low	% Solid:	84

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.0088	U	1	0.038		0.29	mg/Kg	02/12/2018	02/12/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	2/9/2018 12:00:00 AM
Project:	LB93231	Date Received:	2/9/2018 12:00:00 AM
Client Sample ID:	TP-2	SDG No.:	LB93231
Lab Sample ID:	J1517-15	Matrix:	Solid
Level (low/med):	low	% Solid:	78.6

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.013	U	1	0.04		0.3	mg/Kg	02/12/2018	02/12/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	2/9/2018 12:00:00 AM
Project:	LB93231	Date Received:	2/9/2018 12:00:00 AM
Client Sample ID:	OILY-STONES	SDG No.:	LB93231
Lab Sample ID:	J1520-01	Matrix:	Solid
Level (low/med):	low	% Solid:	100

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.0083	U	1	0.033		0.25	mg/Kg	02/12/2018	02/12/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	2/9/2018 12:00:00 AM
Project:	LB93231	Date Received:	2/9/2018 12:00:00 AM
Client Sample ID:	SAMPLE-1	SDG No.:	LB93231
Lab Sample ID:	J1525-01	Matrix:	Solid
Level (low/med):	low	% Solid:	83.1

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.31		1	0.04		0.3	mg/Kg	02/12/2018	02/12/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	2/9/2018 12:00:00 AM
Project:	LB93231	Date Received:	2/9/2018 12:00:00 AM
Client Sample ID:	SAMPLE-2	SDG No.:	LB93231
Lab Sample ID:	J1525-03	Matrix:	Solid
Level (low/med):	low	% Solid:	84.2

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.17	J	1	0.037		0.28	mg/Kg	02/12/2018	02/12/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	2/9/2018 12:00:00 AM
Project:	LB93231	Date Received:	2/9/2018 12:00:00 AM
Client Sample ID:	SAMPLE-3	SDG No.:	LB93231
Lab Sample ID:	J1525-05	Matrix:	Solid
Level (low/med):	low	% Solid:	87.8

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.1	J	1	0.036		0.28	mg/Kg	02/12/2018	02/12/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	2/9/2018 12:00:00 AM
Project:	LB93231	Date Received:	2/9/2018 12:00:00 AM
Client Sample ID:	RO-1-2	SDG No.:	LB93231
Lab Sample ID:	J1526-01	Matrix:	Solid
Level (low/med):	low	% Solid:	90.6

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.035	U	1	0.036		0.27	mg/Kg	02/12/2018	02/12/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	2/9/2018 12:00:00 AM
Project:	LB93231	Date Received:	2/9/2018 12:00:00 AM
Client Sample ID:	OILY-WATER	SDG No.:	LB93231
Lab Sample ID:	J1527-01	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.0019	U	1	0.003		0.005	mg/L	02/12/2018	02/12/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB93231	Date Received:	
Client Sample ID:	LOWPB106455	SDG No.:	LB93231
Lab Sample ID:	LOWPB106455	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.01		1	0.003		0.005	mg/L		02/12/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB93231	Date Received:	
Client Sample ID:	LOWPB106455	SDG No.:	LB93231
Lab Sample ID:	LOWPB106455	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
-----	-----------	-------	------	----	-----	-----	------------	-------	-----------	-----------	----------

Color Before:	Clarity Before:	Texture:
Color After:	Clarity After:	Artifacts:
Comments:		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB93231

SDG No.: LB93231

Contract: _____

Lab Code: CHEM

Case No.: LB93231

SAS No.: LB93231

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	Qual	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV1	Cyanide	0.1	0.099	101		85 - 115		02/12/2018	15:24	LB93231



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB93231

SDG No.: LB93231

Contract: _____ Lab Code: CHEM

Case No.: LB93231

SAS No.: LB93231

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	Qual	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV1	Cyanide	0.26	0.25	104		90 - 110		02/12/2018	15:24	LB93231
CCV2	Cyanide	0.24	0.25	96		90 - 110		02/12/2018	15:44	LB93231
CCV3	Cyanide	0.24	0.25	96		90 - 110		02/12/2018	15:58	LB93231
CCV4	Cyanide	0.25	0.25	100		90 - 110		02/12/2018	15:58	LB93231



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB93231 **SDG No.:** LB93231
Contract: **Lab Code:** CHEM **Case No.:** LB93231 **SAS No.:** LB93231

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB1	Cyanide	-0.00045	+/-0.005	U		0.005		02/12/2018	15:24	LB93231

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB1	Cyanide	-0.00033	+/-0.005	U		0.005		02/12/2018	15:24	LB93231
CCB2	Cyanide	-0.00049	+/-0.005	U		0.005		02/12/2018	15:52	LB93231
CCB3	Cyanide	-0.0001	+/-0.005	U		0.005		02/12/2018	15:58	LB93231
CCB4	Cyanide	-0.00017	+/-0.005	U		0.005		02/12/2018	15:58	LB93231



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB93231

SDG No.: LB93231

Contract: Lab Code: CHEM

Case No.: LB93231

SAS No.: LB93231

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	-----	------	---	------------------	------------------	---------------



GENCHEM
- 3b -
PREPARATION BLANK SUMMARY

Client: LB93231

SDG No.: LB93231

Instrument: Konelab 20

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	LOD mg/Kg	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB106455BL	Cyanide	SOLID		Batch Number:		PB106455		Prep Date:	02/12/2018	
		-0.12	<0.25	U		0.25		02/12/2018	15:44	LB93231
Sample ID	Analyte	Result (mg/L)	Acceptance Limit	Conc Qual	LOD mg/L	CRQL mg/L	M	Analysis Date	Analysis Time	Run
PB106470BL	Cyanide	WATER		Batch Number:		PB106470		Prep Date:	02/12/2018	
		-0.00033	<0.005	U		0.005		02/12/2018	15:52	LB93231



GENCHEM
- 5a -
MATRIX SPIKE SUMMARY

client: LB93231 level: low sdg no.: LB93231
contract: lab code: CHEM case no.: LB93231 sas no.: LB93231
matrix: Solid sample id: J1516-01 client id: WAVE-1MS
Percent Solids for Sample: 81.1 Spiked ID: J1516-01MS Percent Solids for Spike Sample: 81.1

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125	2.1		0.04	U	2.4	87.5		



GENCHEM
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: LB93231 level: low sdg no.: LB93231
contract: lab code: CHEM case no.: LB93231 sas no.: LB93231
matrix: Solid sample id: J1516-01 client id: WAVE-1MSD
Percent Solids for Sample: 81.1 Spiked ID: J1516-01MSD Percent Solids for Spike Sample: 81.1

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125	2.2		0.04	U	2.4	91.7		



GENCHEM
- 5a -
MATRIX SPIKE SUMMARY

client: LB93231 level: low sdg no.: LB93231
contract: lab code: CHEM case no.: LB93231 sas no.: LB93231
matrix: WATER sample id: J1527-01 client id: OILY-WATERMS
Percent Solids for Sample: 0 Spiked ID: J1527-01MS Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/L	75 - 125	0.033		0.003	U	0.04	82.5		



GENCHEM
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: LB93231 level: low sdg no.: LB93231
contract: lab code: CHEM case no.: LB93231 sas no.: LB93231
matrix: WATER sample id: J1527-01 client id: OILY-WATERMSD
Percent Solids for Sample: 0 Spiked ID: J1527-01MSD Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/L	75 - 125	0.034		0.003	U	0.04	85		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB93231 **Level:** LOW **SDG No.:** LB93231
Contract: **Lab Code:** CHEM **Case No.:** LB93231 **SAS No.:** LB93231
Matrix: Solid **Sample ID:** J1516-01 **Client ID:** WAVE-1DUP
Percent Solids for Sample: 81.1 **Duplicate ID** J1516-01DUP **Percent Solids for Spike Sample:** 81.1

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	0.04	U	-0.00082	U			

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB93231 **Level:** LOW **SDG No.:** LB93231
Contract: **Lab Code:** CHEM **Case No.:** LB93231 **SAS No.:** LB93231
Matrix: Solid **Sample ID:** J1516-01MS **Client ID:** WAVE-1MSD
Percent Solids for Sample: 81.1 **Duplicate ID** J1516-01MSD **Percent Solids for Spike Sample:** 81.1

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	2.1		2.2		5		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB93231 **Level:** LOW **SDG No.:** LB93231
Contract: **Lab Code:** CHEM **Case No.:** LB93231 **SAS No.:** LB93231
Matrix: WATER **Sample ID:** J1527-01 **Client ID:** OILY-WATERDUP
Percent Solids for Sample: 0 **Duplicate ID** J1527-01DUP **Percent Solids for Spike Sample:** 0

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/L	20	0.003	U	0.0019	U			

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB93231 **Level:** LOW **SDG No.:** LB93231
Contract: **Lab Code:** CHEM **Case No.:** LB93231 **SAS No.:** LB93231
Matrix: WATER **Sample ID:** J1527-01MS **Client ID:** OILY-WATERMSD
Percent Solids for Sample: 0 **Duplicate ID** J1527-01MSD **Percent Solids for Spike Sample:** 0

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/L	20	0.033		0.034		3		



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB93231

SDG No.: LB93231

Contract:

Lab Code: CHEM

Case No.: LB93231

SAS No.: LB93231

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB106455BS Cyanide	mg/Kg	4.8	4.7		97.9	85 - 115	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB93231

SDG No.: LB93231

Contract:

Lab Code: CHEM

Case No.: LB93231

SAS No.: LB93231

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB106470BS Cyanide	mg/L	0.1	0.1		100	85 - 115	