



## Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB93232

Ambient Barometric Pressure (mmHg): 760

| Reagent/Standard | Lot/Log # |
|------------------|-----------|
| p-xylene (ICV)   | W2168     |

| Seq | LabID       | True Value °F | Dilution | Celsius °C | Result °F | Anal Date  | Anal Time |
|-----|-------------|---------------|----------|------------|-----------|------------|-----------|
| 1   | ICV         | 81            | 1        | 27.0       | 80.6      | 02/12/2018 | 09:02     |
| 2   | J1527-02    |               | 1        | >100.0     | >212.0    | 02/12/2018 | 09:32     |
| 3   | J1527-02DUP |               | 1        | >100.0     | >212.0    | 02/12/2018 | 09:55     |

Result = (Celsius \* 1.8) + 32

LB93232

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP21218

WorkList ID : 108456

Date : 2/12/2018 8:13:05 AM

| Due Date   | Matrix | Sample     | Test        | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|------------|-------------|--------------|----------|------------------|-----------------|--------------|--------|
| 02/14/2018 | Water  | J1527-02 J | Flash Point | Cool 4 deg C | PSEG01   | P52              | OILY-WATER      | 02/09/2018   | 1010A  |

Date/Time 2/12/18 8:30 AM  
Received by: KS  
Relinquished by: CP

Date/Time 2/12/18 10:00 AM  
Received by: CP  
Relinquished by: KS