



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB93237

Temp.: 22.6 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/12/2018	17:00
2	CAL2	1	Water			7.01	02/12/2018	17:01
3	CAL3	1	Water			10.03	02/12/2018	17:03
4	ICV	1	Water			7.00	02/12/2018	17:05
5	CCV1	1	Water			2.02	02/12/2018	17:06
6	J1397-02	1	Solid	20.02	20	8.39	02/12/2018	17:08
7	J1397-02DUP	1	Solid	20.04	20	8.40	02/12/2018	17:10
8	J1397-04	1	Solid	20.01	20	7.51	02/12/2018	17:11
9	J1397-06	1	Solid	20.04	20	9.18	02/12/2018	17:12
10	J1397-08	1	Solid	20.03	20	7.48	02/12/2018	17:15
11	J1398-02	1	Solid	20.05	20	7.11	02/12/2018	17:16
12	J1403-04	1	Solid	20.05	20	9.96	02/12/2018	17:18
13	J1537-07	1	Solid	20.02	20	8.07	02/12/2018	17:22
14	J1537-08	1	Solid	20.04	20	9.43	02/12/2018	17:25
15	J1537-09	1	Solid	20.05	20	9.36	02/12/2018	17:26
16	CCV2	1	Water			12.01	02/12/2018	17:30

230323F

WORKLIST(Hardcopy Internal Chain)

WorkList Name: CORROSIONITY 11403

WorkList ID: 108489

Date: 2/12/2018 4:24:40 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/12/2018	Solid	J1397-02	Corrosivity	Cool 4 deg C	PSEG05	P52	CL-01-021218-A	02/12/2018	9045C
02/12/2018	Solid	J1397-04	Corrosivity	Cool 4 deg C	PSEG05	P52	CL-01-021218-B	02/12/2018	9045C
02/12/2018	Solid	J1397-06	Corrosivity	Cool 4 deg C	PSEG05	P52	CL-01-021218-C	02/12/2018	9045C
02/12/2018	Solid	J1397-08	Corrosivity	Cool 4 deg C	PSEG05	P52	CL-01-021218-D	02/12/2018	9045C
02/12/2018	Solid	J1398-02	Corrosivity	Cool 4 deg C	PSEG05	P52	OK-02-021218	02/12/2018	9045C
02/12/2018	Solid	J1403-04	Corrosivity	Cool 4 deg C	PSEG05	P52	SU-03-021218	02/12/2018	9045C
02/09/2018	Solid	J1537-07	Corrosivity	Cool 4 deg C	LIRO01	P41	SB-01-COMP	02/09/2018	9045C
02/09/2018	Solid	J1537-08	Corrosivity	Cool 4 deg C	LIRO01	P41	SB-02-COMP	02/09/2018	9045C
02/09/2018	Solid	J1537-09	Corrosivity	Cool 4 deg C	LIRO01	P41	SB-03-COMP	02/09/2018	9045C

Date/Time 02-12-17 4:25 PM

Received by: JF

Relinquished by: CR

Date/Time 02-12-17 6:00 PM

Received by: JF

Relinquished by: JF