


7471B

LB93238
INSTRUMENT ID: CV1

Linear

 μ Abs.:

43650



A= 0.0000e+000

B= 2.2712e-004

C= 3.5799e-004

Rho= 0.9996466

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.012	-0.012	-54	0.000	-54				
0.2	0.200	0.186	-0.014	819	0.0 %	819				
2.5	2.500	2.405	-0.095	10588	0.0 %	10588				
5.0	5.000	5.203	0.203	22905	0.0 %	22905				
7.5	7.500	7.503	0.003	33035	0.0 %	33035				
10	10.000	9.914	-0.086	43650	0.0 %	43650				

LB93238 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 12 Feb 2018 18:05:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-54	-	0.0000	7471B	PPB	12 Feb 2018 18:06:48	S	Std
0.2 - 1	50.2	819	-	0.2000	7471B	PPB	12 Feb 2018 18:09:05	S	Std
2.5 - 1	52.5	10588	-	2.5000	7471B	PPB	12 Feb 2018 18:11:23	S	Std
5.0 - 1	55	22905	-	5.0000	7471B	PPB	12 Feb 2018 18:13:40	S	Std
7.5 - 1	57.5	33035	-	7.5000	7471B	PPB	12 Feb 2018 18:19:12	S	Std
10 - 1	510	43650	-	10.0000	7471B	PPB	12 Feb 2018 18:24:04	S	Std
ICV96 - 1	ICV96	17030	3.8683	-7471B	PPB	PPB	12 Feb 2018 18:26:50	U	SMPL
ICB96 - 1	ICB96	-206	-0.0464	-7471B	PPB	PPB	12 Feb 2018 18:29:05	U	SMPL
CCV29 - 1	CCV29	22948	5.2124	-7471B	PPB	PPB	12 Feb 2018 18:31:21	U	SMPL
CCB29 - 1	CCB29	-196	-0.0442	-7471B	PPB	PPB	12 Feb 2018 18:33:37	U	SMPL
CRA - 1	CRA	815	0.1855	-7471B	PPB	PPB	12 Feb 2018 18:35:53	U	SMPL
HighStd - 1	HighStd	43875	9.9654	-7471B	PPB	PPB	12 Feb 2018 18:38:09	U	SMPL
ChkStd - 1	ChkStd	29918	6.7955	-7471B	PPB	PPB	12 Feb 2018 18:40:25	U	SMPL
PB106451BL - 1	PBS	-173	-0.0389	-7471B	PPB	PPB	12 Feb 2018 18:42:43	U	SMPL
PB106451BS - 1	LCSS	16329	3.7091	-7471B	PPB	PPB	12 Feb 2018 18:45:01	U	SMPL
J1516-01 - 1	WAVE-1	997	0.2268	-7471B	PPB	PPB	12 Feb 2018 18:47:18	U	SMPL
J1517-01 - 1	CO-E	545	0.1241	-7471B	PPB	PPB	12 Feb 2018 18:49:35	U	SMPL
J1517-03 - 1	CO-M	323	0.0737	-7471B	PPB	PPB	12 Feb 2018 18:51:53	U	SMPL
J1517-05 - 1	CO-W	150	0.0344	-7471B	PPB	PPB	12 Feb 2018 18:54:11	U	SMPL
CCV30 - 1	CCV30	22786	5.1756	-7471B	PPB	PPB	12 Feb 2018 18:58:43	U	SMPL
CCB30 - 1	CCB30	-200	-0.0451	-7471B	PPB	PPB	12 Feb 2018 19:00:58	U	SMPL
J1517-07 - 1	TP-1	2510	0.5704	-7471B	PPB	PPB	12 Feb 2018 19:06:18	U	SMPL
J1517-07DUP - 1	TP-1DUP	2439	0.5543	-7471B	PPB	PPB	12 Feb 2018 19:08:34	U	SMPL
J1517-07MS - 1	TP-1MS	17211	3.9094	-7471B	PPB	PPB	12 Feb 2018 19:10:50	U	SMPL
J1517-07MSD - 1	TP-1MSD	17176	3.9014	-7471B	PPB	PPB	12 Feb 2018 19:13:07	U	SMPL
J1517-15 - 1	TP-2	1419	0.3226	-7471B	PPB	PPB	12 Feb 2018 19:15:24	U	SMPL
J1520-01 - 1	OILY-STONES	627	0.1428	-7471B	PPB	PPB	12 Feb 2018 19:17:40	U	SMPL
J1520-04 - 1	OILY-DEBRIS	-2	-0.0001	-7471B	PPB	PPB	12 Feb 2018 19:19:58	U	SMPL
J1525-01 - 1	SAMPLE-1	29798	6.7682	-7471B	PPB	PPB	12 Feb 2018 19:22:15	U	SMPL
J1525-03 - 1	SAMPLE-2	7191	1.6336	-7471B	PPB	PPB	12 Feb 2018 19:24:33	U	SMPL
J1525-05 - 1	SAMPLE-3	17754	4.0327	-7471B	PPB	PPB	12 Feb 2018 19:26:50	U	SMPL
CCV31 - 1	CCV31	21387	4.8579	-7471B	PPB	PPB	12 Feb 2018 19:29:07	U	SMPL
CCB31 - 1	CCB31	-42	-0.0092	-7471B	PPB	PPB	12 Feb 2018 19:31:23	U	SMPL
J1526-01 - 1	RO-1-2	1922	0.4369	-7471B	PPB	PPB	12 Feb 2018 19:33:40	U	SMPL
J1517-07LX5 - 1		531	0.1210	-7471B	PPB	PPB	12 Feb 2018 19:35:56	U	SMPL
J1517-07A - 1		18677	4.2424	-7471B	PPB	PPB	12 Feb 2018 19:38:12	U	SMPL
CCV32 - 1	CCV32	21030	4.7768	-7471B	PPB	PPB	12 Feb 2018 19:40:29	U	SMPL
CCB32 - 1	CCB32	-39	-0.0085	-7471B	PPB	PPB	12 Feb 2018 19:42:46	U	SMPL