

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus, TotalSUPERVISOR REVIEW BY: apatelRun Number: LB93251BALANCE ID: WC SC-4

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP61786
calibration std. phosphate 1 ppm	WP61846
calibration std. phosphate 0.5 ppm	WP61845
calibration std. phosphate 0.3 ppm	WP61844
calibration std. 0 ppm	WP61841
phosphate ICV std.	WP61848
phosphate CCV std.	WP61847
calibration std. phosphate 0.1 ppm	WP61843
calibration std. phosphate 0.05 ppm	WP61842
Combined reagent	WP61853
11 N sulfuric acid	WP61601
Phenolphthalein indicator	WP61777
Sodium hydroxide, 1N	WP61776

Intercept: -0.0007Slope: 0.6498Regression: 0.999842

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	02/13/2018	11:40
2	CAL2	0.05	1	50	50	0.033	02/13/2018	11:40
3	CAL3	0.10	1	50	50	0.066	02/13/2018	11:41
4	CAL4	0.30	1	50	50	0.186	02/13/2018	11:41
5	CAL5	0.50	1	50	50	0.329	02/13/2018	11:42
6	CAL6	1.00	1	50	50	0.649	02/13/2018	11:43



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB93251

BALANCE ID: WC SC-4

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.318	0.490	02/13/2018	11:43
2	ICB		1	50	50	0.000	0.001	02/13/2018	11:44
3	CCV1	0.50	1	50	50	0.325	0.501	02/13/2018	11:44
4	CCB1		1	50	50	0.000	0.001	02/13/2018	11:45
5	PB106516BL		1	1.00	50	0.000	0.001	02/13/2018	11:45
6	PB106516BS	0.50	1	1.01	50	0.317	0.489	02/13/2018	11:51
7	J1161-01		1	1.00	50	0.019	0.030	02/13/2018	11:51
8	J1161-02		1	1.01	50	0.071	0.110	02/13/2018	11:52
9	J1161-03		1	1.00	50	0.023	0.036	02/13/2018	11:52
10	CCV2	0.50	1	50	50	0.326	0.503	02/13/2018	11:53
11	CCB2		1	50	50	0.000	0.001	02/13/2018	11:54