



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB93262

Temp.: 23.2 °C

SLOPE.: 99.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/13/2018	17:35
2	CAL2	1	Water			6.99	02/13/2018	17:37
3	CAL3	1	Water			10.01	02/13/2018	17:39
4	ICV	1	Water			7.00	02/13/2018	17:41
5	CCV1	1	Water			2.00	02/13/2018	17:43
6	J1549-01	1	Water			9.04	02/13/2018	17:45
7	J1549-01DUP	1	Water			9.05	02/13/2018	17:47
8	J1549-02	1	Water			11.64	02/13/2018	17:49
9	CCV2	1	Water			12.00	02/13/2018	17:50

WORKLIST(Hardcopy Internal Chain)

LB 93262

WorkList Name : CORROSIVITY J1549

WorkList ID : 108527

Date : 2/13/2018 5:22:25 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/13/2018	Water	J1549-01	Corrosivity	Cool 4 deg C	PSEG01	P53	Frac-Tank-N37475	02/13/2018	9040C
02/13/2018	Water	J1549-02	Corrosivity	Cool 4 deg C	PSEG01	P53	Frac-Tank-N28427	02/13/2018	9040C

Date/Time 02/13/18 5:25 PM

Received by: SP

Relinquished by: SP

Date/Time 02/13/18 6:00 PM

Received by: SP

Relinquished by: SP