


7470A

L393269
INSTRUMENT ID: CV2

Linear

 μ Abs.:

37366



A= 0.0000e+000

B= 2.7054e-004

C= -7.0451e-002

Rho= 0.9999688

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.035	0.035	389	0.000	389				
0.2	0.200	0.201	0.001	1004	0.0 %	1004				
2.5	2.500	2.488	-0.012	9458	0.0 %	9458				
5.0	5.000	4.956	-0.044	18580	0.0 %	18580				
7.5	7.500	7.481	-0.019	27914	0.0 %	27914				
10.0	10.000	10.038	0.038	37366	0.0 %	37366				

LB93269 INSTRUMENT : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Feb 2018 10:17:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	389	-	0.0000	7470A	PPB	14 Feb 2018 10:26:09	S	Std
0.2 - 1	50.2	1004	-	0.2000	7470A	PPB	14 Feb 2018 10:28:18	S	Std
2.5 - 1	52.5	9458	-	2.5000	7470A	PPB	14 Feb 2018 10:30:26	S	Std
5.0 - 1	55	18580	-	5.0000	7470A	PPB	14 Feb 2018 10:32:34	S	Std
7.5 - 1	57.5	27914	-	7.5000	7470A	PPB	14 Feb 2018 10:34:42	S	Std
10.0 - 1	510	37366	-	10.0000	7470A	PPB	14 Feb 2018 10:36:51	S	Std
ICV98 - 1	ICV98	14568	3.8707	-	7470A	PPB	14 Feb 2018 10:39:28	U	SMPL
ICB98 - 1	ICB98	261	0.0002	-	7470A	PPB	14 Feb 2018 10:41:34	U	SMPL
CCV38 - 1	CCV38	18931	5.0511	-	7470A	PPB	14 Feb 2018 10:43:41	U	SMPL
CCB38 - 1	CCB38	258	-0.0007	-	7470A	PPB	14 Feb 2018 10:45:47	U	SMPL
CRA - 1	CRA	1028	0.2077	-	7470A	PPB	14 Feb 2018 10:47:55	U	SMPL
HighStd - 1	HighStd	37002	9.9399	-	7470A	PPB	14 Feb 2018 10:50:01	U	SMPL
ChkStd - 1	ChkStd	25345	6.7863	-	7470A	PPB	14 Feb 2018 10:52:08	U	SMPL
PB106521BL - 1	PBW	138	-0.0331	-	7470A	PPB	14 Feb 2018 10:59:09	U	SMPL
PB106521BS - 1	LCSW	14018	3.7219	-	7470A	PPB	14 Feb 2018 11:01:15	U	SMPL
PB106521TB - 1	PB106495TB	-52	-0.0845	-	7470A	PPB	14 Feb 2018 11:05:49	U	SMPL
J1397-02 - 1	CL-01-021218-A	-55	-0.0853	-	7470A	PPB	14 Feb 2018 11:07:56	U	SMPL
J1397-02DUP - 1	CL-01-021218-ADUP	-77	-0.0913	-	7470A	PPB	14 Feb 2018 11:10:05	U	SMPL
J1397-02MS - 1	CL-01-021218-AMS	13493	3.5799	-	7470A	PPB	14 Feb 2018 11:12:13	U	SMPL
J1397-02MSD - 1	CL-01-021218-AMSD	13827	3.6702	-	7470A	PPB	14 Feb 2018 11:14:19	U	SMPL
CCV39 - 1	CCV39	18715	4.9926	-	7470A	PPB	14 Feb 2018 11:16:26	U	SMPL
CCB39 - 1	CCB39	225	-0.0096	-	7470A	PPB	14 Feb 2018 11:18:33	U	SMPL
J1397-04 - 1	CL-01-021218-B	-53	-0.0848	-	7470A	PPB	14 Feb 2018 11:20:40	U	SMPL
J1397-06 - 1	CL-01-021218-C	-34	-0.0796	-	7470A	PPB	14 Feb 2018 11:22:47	U	SMPL
J1397-08 - 1	CL-01-021218-D	-42	-0.0818	-	7470A	PPB	14 Feb 2018 11:24:54	U	SMPL
J1398-02 - 1	OK-02-021218	-50	-0.0840	-	7470A	PPB	14 Feb 2018 11:27:02	U	SMPL
J1403-04 - 1	SU-03-021218	-34	-0.0796	-	7470A	PPB	14 Feb 2018 11:29:09	U	SMPL
J1537-07 - 1	SB-01-COMP	-27	-0.0778	-	7470A	PPB	14 Feb 2018 11:31:17	U	SMPL
J1537-08 - 1	SB-02-COMP	6	-0.0688	-	7470A	PPB	14 Feb 2018 11:33:25	U	SMPL
J1537-09 - 1	SB-03-COMP	37	-0.0604	-	7470A	PPB	14 Feb 2018 11:35:33	U	SMPL
J1397-02LX5 - 1		286	0.0069	-	7470A	PPB	14 Feb 2018 11:46:21	U	SMPL
J1397-02A - 1		14467	3.8434	-	7470A	PPB	14 Feb 2018 11:48:28	U	SMPL
CCV40 - 1	CCV40	18482	4.9296	-	7470A	PPB	14 Feb 2018 11:50:35	U	SMPL
CCB40 - 1	CCB40	243	-0.0047	-	7470A	PPB	14 Feb 2018 11:52:42	U	SMPL