

LB93275

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

2/14/2018 11:35

Test: Phenolics-mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.013	0.0	0.114	
ICB1	0.002	0.0	0.009	
CCV1	0.999	0.0	0.112	
CCB1	-0.001	0.0	0.009	
PB106537BL	0.002	0.0	0.009	
PB106537BS	0.997	0.0	0.112	
J1161-10	0.015	0.0	0.010	
J1161-12	0.015	0.0	0.010	
CCV2	1.000	0.0	0.112	
CCB2	0.003	0.0	0.009	

N	10
Mean	0.405
SD	0.5145
CV%	127.16

AS

Aquakem v. 7.2AQ1

Results from time period:

Wed Feb 14 11:14:36 2018

Wed Feb 14 11:34:10 2018

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0428	mg/l	2/14/2018 9:50:28	
0.1PPM	A	Phenolics-	P	0.0981	mg/l	2/14/2018 9:50:29	
0.25PPM	A	Phenolics-	P	0.251	mg/l	2/14/2018 9:50:30	
0.50PPM	A	Phenolics-	P	0.5082	mg/l	2/14/2018 9:50:31	
1.0PPM	A	Phenolics-	P	1.0037	mg/l	2/14/2018 9:50:32	
2.0PPM	A	Phenolics-	P	1.9963	mg/l	2/14/2018 9:50:33	
ICV1	S	Phenolics-	P	1.0128	mg/l	2/14/2018 11:14:36	
ICB1	S	Phenolics-	P	0.0022	mg/l	2/14/2018 11:14:37	
CCV1	S	Phenolics-	P	0.9992	mg/l	2/14/2018 11:14:38	
CCB1	S	Phenolics-	P	-0.0009	mg/l	2/14/2018 11:14:39	
PB106537BL	S	Phenolics-	P	0.0017	mg/l	2/14/2018 11:34:05	
PB106537BS	S	Phenolics-	P	0.9973	mg/l	2/14/2018 11:34:06	
J1161-10	S	Phenolics-	P	0.0153	mg/l	2/14/2018 11:34:07	
J1161-12	S	Phenolics-	P	0.0154	mg/l	2/14/2018 11:34:08	
CCV2	S	Phenolics-	P	1	mg/l	2/14/2018 11:34:09	
CCB2	S	Phenolics-	P	0.003	mg/l	2/14/2018 11:34:10	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

2/14/2018 10:02


Test Phenolics-

Accepted 2/14/2018 10:02

Factor 9.656
Bias 0.009

Coeff. of det. 0.999946

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.013	0.0428	0.0500	
2	Phenol-2	0.019	0.0981	0.1000	
3	Phenol-2	0.035	0.2510	0.2500	
4	Phenol-2	0.061	0.5082	0.5000	
5	Phenol-2	0.113	1.0037	1.0000	
6	Phenol-2	0.215	1.9963	2.0000	