



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB93277

Reagent/Standard	Lot/Log #
Residual chlorine std, 0.8PPM	WP61923
Residual chlorine std, 0PPM	WP61919
Residual chlorine std, 0.1PPM	WP61920
Residual chlorine LOQ, 0.2PPM	WP61927
Residual chlorine std, 0.2PPM	WP61921
Residual chlorine std, and CCV , 0.4PPM	WP61922
Residual chlorine std, 1.6PPM	WP61924
Residual chlorine ICV-LCS, 0.4PPM	WP61925
Residual chlorine LOD, 0.05PPM	WP61926
Total Chlorine Powder Pillows	W1960

Intercept: 0.0063

Slope: 1.0362

Regression: 0.999954

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	02/14/2018	13:13
2	CAL2	0.1	1	0.000	0.110	0.110	02/14/2018	13:13
3	CAL3	0.2	1	0.000	0.210	0.210	02/14/2018	13:14
4	CAL4	0.4	1	0.000	0.430	0.430	02/14/2018	13:14
5	CAL5	0.8	1	0.000	0.840	0.840	02/14/2018	13:15
6	CAL6	1.6	1	0.000	1.660	1.660	02/14/2018	13:15



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB93277

Seq	Lab ID	Initial Weight	Final Vol	True Value	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.000	0.430	0.430	0.409	02/14/2018	13:16
2	ICB				1	0.000	0.000	0.000	-0.006	02/14/2018	13:16
3	CCV1			0.4	1	0.000	0.430	0.430	0.409	02/14/2018	13:17
4	CCB1				1	0.000	0.000	0.000	-0.006	02/14/2018	13:17
5	LB93277BL	50	50		1	0.000	0.000	0.000	-0.006	02/14/2018	13:17
6	LB93277BS	50	50	0.4	1	0.000	0.420	0.420	0.399	02/14/2018	13:18
7	J1161-07	50	50		1	0.000	0.070	0.070	0.061	02/14/2018	13:18
8	J1161-08	50	50		1	0.000	0.220	0.220	0.206	02/14/2018	13:19
9	J1161-09	50	50		1	0.000	0.060	0.060	0.052	02/14/2018	13:19
10	CCV2			0.4	1	0.000	0.400	0.400	0.380	02/14/2018	13:20
11	CCB2				1	0.000	0.000	0.000	-0.006	02/14/2018	13:20