


ISM02.4

LB 93278
INSTRUMENT ID: CY1

Linear

 μ Abs.:

43738



A= 0.0000e+000

B= 2.2976e-004

C= -2.6084e-002

Rho= 0.9999361

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.036	-0.036	-43	0.000	-43					-
0.2	0.200	0.226	0.026	1098	0.0 %	1098					13
2.5	2.500	2.561	0.061	11261	0.0 %	11261					2
5.0	5.000	4.936	-0.064	21598	0.0 %	21598					-1
7.5	7.500	7.489	-0.011	32711	0.0 %	32711					0
10.0	10.000	10.023	0.023	43738	0.0 %	43738					0

123
2114118

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	-43	-		0 ISM02.4	PPB	2/14/2018 14:02	Std
	0.2 S0.2	1098	-		0.2 ISM02.4	PPB	2/14/2018 14:09	Std
	2.5 S2.5	11261	-		2.5 ISM02.4	PPB	2/14/2018 14:11	Std
	5 S5	21598	-		5 ISM02.4	PPB	2/14/2018 14:14	Std
	7.5 S7.5	32711	-		7.5 ISM02.4	PPB	2/14/2018 14:16	Std
	10 S10	43738	-		10 ISM02.4	PPB	2/14/2018 14:18	Std
ICV	ICV	16930	3.8637 -		ISM02.4	PPB	2/14/2018 14:21	SMPL
ICB	ICB	-59	-0.0396 -		ISM02.4	PPB	2/14/2018 14:23	SMPL
CCV041	CCV041	21515	4.9171 -		ISM02.4	PPB	2/14/2018 14:25	SMPL
CCB041	CCB041	-77	-0.0438 -		ISM02.4	PPB	2/14/2018 14:28	SMPL
PB106545BL	PBW003	49	-0.0148 -		ISM02.4	PPB	2/14/2018 14:30	SMPL
J1542-01	MCOAA0	-227	-0.0782 -		ISM02.4	PPB	2/14/2018 14:32	SMPL
J1542-01DUP	MCOAA0DUP	-242	-0.0817 -		ISM02.4	PPB	2/14/2018 14:34	SMPL
J1542-01MS	MCOAA0MS	4015	0.8964 -		ISM02.4	PPB	2/14/2018 14:37	SMPL
J1542-02	MCOAA1	-280	-0.0904 -		ISM02.4	PPB	2/14/2018 14:39	SMPL
J1542-03	MCOAA2	-272	-0.0886 -		ISM02.4	PPB	2/14/2018 14:41	SMPL
J1542-04	MCOAA3	-271	-0.0883 -		ISM02.4	PPB	2/14/2018 14:44	SMPL
CCV042	CCV042	21381	4.8863 -		ISM02.4	PPB	2/14/2018 14:46	SMPL
CCB042	CCB042	-94	-0.0477 -		ISM02.4	PPB	2/14/2018 14:48	SMPL