


7471B

LB 93285
INSTRUMENT ID: CV1

Linear

 μ Abs.:

44614



A= 0.0000e+000

B= 2.2561e-004

C= -2.9334e-002

Rho= 0.9999139

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.022	-0.022	34	0.000	34				
0.2	0.200	0.204	0.004	1034	0.0 %	1034				
2.5	2.500	2.520	0.020	11298	0.0 %	11298				
5.0	5.000	5.056	0.056	22540	0.0 %	22540				
7.5	7.500	7.406	-0.094	32958	0.0 %	32958				
10	10.000	10.036	0.036	44614	0.0 %	44614				

LB93285 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Feb 2018 15:42:43

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
J0 - 1	50	34	-	0.0000	7471B	PPB	14 Feb 2018 15:48:28	S	Std
J2 - 1	50.2	1034	-	0.2000	7471B	PPB	14 Feb 2018 15:56:08	S	Std
J2.5 - 1	52.5	11298	-	2.5000	7471B	PPB	14 Feb 2018 15:58:26	S	Std
J5.0 - 1	55	22540	-	5.0000	7471B	PPB	14 Feb 2018 16:00:43	S	Std
J7.5 - 1	57.5	32958	-	7.5000	7471B	PPB	14 Feb 2018 16:03:00	S	Std
J10 - 1	510	44614	-	10.0000	7471B	PPB	14 Feb 2018 16:05:17	S	Std
ICV99 - 1	ICV99	16328	3.6544	-	7471B	PPB	14 Feb 2018 16:07:58	U	SMPL
ICB99 - 1	ICB99	-98	-0.0514	-	7471B	PPB	14 Feb 2018 16:10:13	U	SMPL
CCV43 - 1	CCV43	22414	5.0275	-	7471B	PPB	14 Feb 2018 16:12:29	U	SMPL
CCB43 - 1	CCB43	-125	-0.0575	-	7471B	PPB	14 Feb 2018 16:14:45	U	SMPL
CRA - 1	CRA	1052	0.2080	-	7471B	PPB	14 Feb 2018 16:17:00	U	SMPL
HighStd - 1	HighStd	44670	10.0486	-	7471B	PPB	14 Feb 2018 16:19:16	U	SMPL
ChkStd - 1	ChkStd	30850	6.9307	-	7471B	PPB	14 Feb 2018 16:21:31	U	SMPL
PB106532BL - 1	PBS	-91	-0.0499	-	7471B	PPB	14 Feb 2018 16:23:49	U	SMPL
PB106532BS - 1	LCSS	14812	3.3124	-	7471B	PPB	14 Feb 2018 16:26:08	U	SMPL
J1397-01 - 1	CL-01-021218-A	5350	1.1777	-	7471B	PPB	14 Feb 2018 16:28:25	U	SMPL
J1397-03 - 1	CL-01-021218-B	1613	0.3346	-	7471B	PPB	14 Feb 2018 16:30:42	U	SMPL
J1397-05 - 1	CL-01-021218-C	725	0.1342	-	7471B	PPB	14 Feb 2018 16:33:00	U	SMPL
J1397-07 - 1	CL-01-021218-D	1350	0.2752	-	7471B	PPB	14 Feb 2018 16:35:18	U	SMPL
J1398-01 - 1	OK-02-021218	4398	0.9629	-	7471B	PPB	14 Feb 2018 16:37:34	U	SMPL
CCV44 - 1	CCV44	22325	5.0074	-	7471B	PPB	14 Feb 2018 16:39:50	U	SMPL
CCB44 - 1	CCB44	-105	-0.0530	-	7471B	PPB	14 Feb 2018 16:42:06	U	SMPL
J1403-03 - 1	SU-03-021218	594	0.1047	-	7471B	PPB	14 Feb 2018 16:44:22	U	SMPL
J1530-45 - 1	WC-01	4601	1.0087	-	7471B	PPB	14 Feb 2018 16:46:38	U	SMPL
J1530-46 - 1	WC-02	1348	0.2748	-	7471B	PPB	14 Feb 2018 16:48:54	U	SMPL
J1530-47 - 1	WC-03	167	0.0083	-	7471B	PPB	14 Feb 2018 16:51:10	U	SMPL
J1530-48 - 1	WC-04	868	0.1665	-	7471B	PPB	14 Feb 2018 16:53:27	U	SMPL
J1540-01 - 1	CB-1-NW(10-15)	46632	10.4912	-	7471B	PPB	14 Feb 2018 16:55:43	U	SMPL
J1540-02 - 1	CB-3-LINE-NW@3	4806	1.0549	-	7471B	PPB	14 Feb 2018 16:58:01	U	SMPL
J1540-03 - 1	CB-3-LINE-SW@3	5035	1.1066	-	7471B	PPB	14 Feb 2018 17:00:20	U	SMPL
J1540-04 - 1	CB-4-LINE-NW@2.5	1406	0.2879	-	7471B	PPB	14 Feb 2018 17:02:38	U	SMPL
J1540-05 - 1	CB-4-LINE-SW@2.5	1147	0.2294	-	7471B	PPB	14 Feb 2018 17:04:54	U	SMPL
CCV45 - 1	CCV45	22346	5.0121	-	7471B	PPB	14 Feb 2018 17:07:10	U	SMPL
CCB45 - 1	CCB45	-101	-0.0521	-	7471B	PPB	14 Feb 2018 17:09:27	U	SMPL
J1540-06 - 1	CB-5-LINE-NW@3	1897	0.3986	-	7471B	PPB	14 Feb 2018 17:11:43	U	SMPL
J1540-07 - 1	CB-5-LINE-SW@3	1845	0.3869	-	7471B	PPB	14 Feb 2018 17:13:59	U	SMPL
J1540-07DUP - 1	CB-5-LINE-SW@3DUP	1903	0.4000	-	7471B	PPB	14 Feb 2018 17:16:16	U	SMPL
J1540-07MS - 1	CB-5-LINE-SW@3MS	18905	4.2358	-	7471B	PPB	14 Feb 2018 17:18:32	U	SMPL
J1540-07MSD - 1	CB-5-LINE-SW@3MSD	17951	4.0206	-	7471B	PPB	14 Feb 2018 17:20:49	U	SMPL
PB106533BL - 1	PBS	-10	-0.0316	-	7471B	PPB	14 Feb 2018 17:23:06	U	SMPL
PB106533BS - 1	LCSS	15317	3.4263	-	7471B	PPB	14 Feb 2018 17:25:23	U	SMPL
J1539-01 - 1	MAIN-ST-LINE-1-WW@4	2592	0.5554	-	7471B	PPB	14 Feb 2018 17:27:41	U	SMPL
J1539-02 - 1	MAIN-ST-LINE-1-EW@4	2433	0.5196	-	7471B	PPB	14 Feb 2018 17:29:59	U	SMPL
J1539-03 - 1	MAIN-ST-LINE-2-WW@4	1893	0.3977	-	7471B	PPB	14 Feb 2018 17:32:15	U	SMPL
CCV46 - 1	CCV46	22075	4.9510	-	7471B	PPB	14 Feb 2018 17:34:31	U	SMPL
CCB46 - 1	CCB46	-90	-0.0496	-	7471B	PPB	14 Feb 2018 17:36:48	U	SMPL
J1539-04 - 1	MAIN-ST-LINE-2-EW@4	1642	0.3411	-	7471B	PPB	14 Feb 2018 17:39:04	U	SMPL
J1539-05 - 1	MAIN-ST-LINE-3-WW@4.5	1483	0.3052	-	7471B	PPB	14 Feb 2018 17:41:20	U	SMPL
J1539-06 - 1	MAIN-ST-LINE-3-EW@4.5	3187	0.6897	-	7471B	PPB	14 Feb 2018 17:43:37	U	SMPL
J1539-07 - 1	12-ST-LATERAL-LINE-1-NW@5	1440	0.2955	-	7471B	PPB	14 Feb 2018 17:45:53	U	SMPL
J1539-08 - 1	12-ST-LATERAL-LINE-1-SW@5	1638	0.3402	-	7471B	PPB	14 Feb 2018 17:48:10	U	SMPL
J1539-09 - 1	4-15-TRENCH-1-NW@5	1791	0.3747	-	7471B	PPB	14 Feb 2018 17:50:27	U	SMPL
J1539-10 - 1	4-15-TRENCH-1-SW@5	4659	1.0218	-	7471B	PPB	14 Feb 2018 17:52:44	U	SMPL
J1539-11 - 1	12-ST-LATERAL-LINE-2-NW@5	3895	0.8494	-	7471B	PPB	14 Feb 2018 17:55:02	U	SMPL
J1539-12 - 1	12-ST-LATERAL-LINE-2-SW@5	5678	1.2517	-	7471B	PPB	14 Feb 2018 17:57:19	U	SMPL
J1539-14 - 1	4-15-TRENCH-2-SW(2-4)	8431	1.8728	-	7471B	PPB	14 Feb 2018 17:59:36	U	SMPL
CCV47 - 1	CCV47	22484	5.0432	-	7471B	PPB	14 Feb 2018 18:01:52	U	SMPL
CCB47 - 1	CCB47	-94	-0.0505	-	7471B	PPB	14 Feb 2018 18:04:08	U	SMPL
J1539-15 - 1	4-15-TRENCH-2-NW(4-6)	2083	0.4406	-	7471B	PPB	14 Feb 2018 18:06:25	U	SMPL
J1539-16 - 1	4-15-TRENCH-2-SW(4-6)	2069	0.4374	-	7471B	PPB	14 Feb 2018 18:08:41	U	SMPL
J1539-17 - 1	CB-1-NW(0-4)	7248	1.6059	-	7471B	PPB	14 Feb 2018 18:10:57	U	SMPL
J1539-18 - 1	CB-1-NW(4-9)	46689	10.5041	-	7471B	PPB	14 Feb 2018 18:13:14	U	SMPL
J1539-19 - 1	CB-1-SW(0-4)	4129	0.9022	-	7471B	PPB	14 Feb 2018 18:15:31	U	SMPL
J1539-20 - 1	CB-1-SW(4-9)	79494	17.9052	-	7471B	PPB	14 Feb 2018 18:17:49	U	SMPL
J1539-13 - 1	4-15-TRENCH-2-NW(2-4)	12249	2.7341	-	7471B	PPB	14 Feb 2018 18:20:06	U	SMPL
J1539-13DUP - 1	4-15-TRENCH-2-NW(2-4)DUP	10742	2.3942	-	7471B	PPB	14 Feb 2018 18:22:26	U	SMPL
J1539-21 - 1	J1539-13MS	28851	6.4797	-	7471B	PPB	14 Feb 2018 18:24:44	U	SMPL
J1539-22 - 1	J1539-13MSD	28307	6.3570	-	7471B	PPB	14 Feb 2018 18:29:48	U	SMPL
CCV48 - 1	CCV48	22184	4.9756	-	7471B	PPB	14 Feb 2018 18:32:05	U	SMPL
CCB48 - 1	CCB48	-130	-0.0587	-	7471B	PPB	14 Feb 2018 18:34:22	U	SMPL
J1540-07LX5 - 1		365	0.0530	-	7471B	PPB	14 Feb 2018 18:36:40	U	SMPL
J1540-07A - 1		18580	4.1625	-	7471B	PPB	14 Feb 2018 18:38:56	U	SMPL
J1539-13LX5 - 1		2591	0.5552	-	7471B	PPB	14 Feb 2018 18:41:13	U	SMPL
J1539-13A - 1		27529	6.1814	-	7471B	PPB	14 Feb 2018 18:43:30	U	SMPL
J1540-01DLX2 - 1		22510	5.0491	-	7471B	PPB	14 Feb 2018 18:45:46	U	SMPL
J1539-18DLX2 - 1		22715	5.0954	-	7471B	PPB	14 Feb 2018 18:48:05	U	SMPL
J1539-20DLX5 - 1		14675	3.2815	-	7471B	PPB	14 Feb 2018 18:50:23	U	SMPL
CCV49 - 1	CCV49	22021	4.9388	-	7471B	PPB	14 Feb 2018 18:52:41	U	SMPL
CCB49 - 1	CCB49	-118	-0.0560	-	7471B	PPB	14 Feb 2018 18:54:59	U	SMPL