



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB93288

Reagent/Standard	Lot/Log #
Residual chlorine std, 0.8PPM	WP61923
Residual chlorine std, 0PPM	WP61919
Residual chlorine std, 0.1PPM	WP61920
Residual chlorine LOQ, 0.2PPM	WP61927
Residual chlorine std, 0.2PPM	WP61921
Residual chlorine std, and CCV , 0.4PPM	WP61922
Residual chlorine std, 1.6PPM	WP61924
Residual chlorine ICV-LCS, 0.4PPM	WP61925
Residual chlorine LOD, 0.05PPM	WP61926
Total Chlorine Powder Pillows	W1960

Intercept: 0.0063

Slope: 1.0362

Regression: 0.999954

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	02/14/2018	15:13
2	CAL2	0.1	1	0.000	0.110	0.110	02/14/2018	15:13
3	CAL3	0.2	1	0.000	0.210	0.210	02/14/2018	15:14
4	CAL4	0.4	1	0.000	0.430	0.430	02/14/2018	15:14
5	CAL5	0.8	1	0.000	0.840	0.840	02/14/2018	15:15
6	CAL6	1.6	1	0.000	1.660	1.660	02/14/2018	15:15



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB93288

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.000	0.430	0.430	0.409	02/14/2018	15:16
2	ICB				1	0.000	0.000	0.000	-0.006	02/14/2018	15:16
3	CCV1			0.4	1	0.000	0.400	0.400	0.380	02/14/2018	15:21
4	CCB1				1	0.000	0.000	0.000	-0.006	02/14/2018	15:21
5	LB93288BL	50	50		1	0.000	0.000	0.000	-0.006	02/14/2018	15:22
6	LB93288BS	50	50	0.4	1	0.000	0.420	0.420	0.399	02/14/2018	15:22
7	J1563-03	50	50		1	0.000	0.000	0.000	-0.006	02/14/2018	15:23
8	J1563-03DUP	50	50		1	0.000	0.000	0.000	-0.006	02/14/2018	15:23
9	CCV2			0.4	1	0.000	0.410	0.410	0.390	02/14/2018	15:24
10	CCB2				1	0.000	0.000	0.000	-0.006	02/14/2018	15:24