



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB93294

Temp.: 24.3 °C

BalanceID: WC SC-4

SLOPE.: 99.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/14/2018	18:10
2	CAL2	1	Water			7.00	02/14/2018	18:12
3	CAL3	1	Water			10.01	02/14/2018	18:14
4	ICV	1	Water			7.00	02/14/2018	18:16
5	CCV1	1	Water			2.00	02/14/2018	18:18
6	J1560-09	1	Solid	20.01	20	7.76	02/14/2018	18:35
7	J1565-02	1	Solid	20.03	20	7.83	02/14/2018	18:37
8	J1565-02DUP	1	Solid	20.02	20	7.84	02/14/2018	18:39
9	J1565-06	1	Solid	20.04	20	6.01	02/14/2018	18:41
10	CCV2	1	Water			12.00	02/14/2018	18:44

WORKLIST(Hardcopy Internal Chain)

1543294

WorkList Name : PH J1561

WorkList ID : 108567

Date : 2/14/2018 6:06:03 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/14/2018	Solid	J1560-09-D	pH	Cool 4 deg C	PSEG02	P63	15228-9	02/14/2018	9045C
02/14/2018	Solid	J1565-02-V	pH	Cool 4 deg C	PSEG02	P62	15340	02/14/2018	9045C
02/14/2018	Solid	J1565-06-F	pH	Cool 4 deg C	PSEG02	P62	19558	02/14/2018	9045C

Date/Time 02.14.18 6:06 PM
 Received by: Jo
 Relinquished by: CB

Date/Time 02-14-18 7:10 PM
 Received by: CB
 Relinquished by: Jo