



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB93295

Temp.: 24.3 °C

SLOPE.: 99.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/14/2018	18:10
2	CAL2	1	Water			7.00	02/14/2018	18:12
3	CAL3	1	Water			10.01	02/14/2018	18:14
4	ICV	1	Water			7.00	02/14/2018	18:16
5	CCV1	1	Water			2.00	02/14/2018	18:18
6	J1565-01	1	Water			6.98	02/14/2018	18:20
7	J1565-01DUP	1	Water			6.99	02/14/2018	18:22
8	J1566-03	1	Water			5.28	02/14/2018	18:24
9	J1566-05	1	Water			5.23	02/14/2018	18:26
10	CCV2	1	Water			12.00	02/14/2018	18:30

WORKLIST(Hardcopy Internal Chain)

15093205

Worklist Name : PH J1560

Worklist ID : 108565

Date : 2/14/2018 6:04:34 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/14/2018	Water	J1560-07-D	pH	Cool 4 deg C	PSEG02	P63	15228-9	02/14/2018	9040C
02/14/2018	Water	J1565-01-C	pH	Cool 4 deg C	PSEG02	P62	18273	02/14/2018	9040C
02/14/2018	Water	J1566-03-Y	pH	Cool 4 deg C	PSEG01	P62	1452-56	02/14/2018	9040C
02/14/2018	Water	J1566-05-Y	pH	Cool 4 deg C	PSEG01	P62	1457-61	02/14/2018	9040C

Date/Time 6-15-18 6:06 PM
 Received by: JP
 Relinquished by: CB

Date/Time 02-14-18 3:10 PM
 Received by: CB
 Relinquished by: JP