



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB93296

Temp.: 23.6 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 2.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/14/2018	17:45
2	CAL2	1	Water			7.01	02/14/2018	17:46
3	CAL3	1	Water			10.03	02/14/2018	17:47
4	ICV	1	Water			7.00	02/14/2018	17:48
5	CCV1	1	Water			2.01	02/14/2018	17:50
6	J1393-02	1	Solid	20.02	20	8.37	02/14/2018	17:51
7	J1393-02DUP	1	Solid	20.03	20	8.38	02/14/2018	18:00
8	J1393-04	1	Solid	20.05	20	9.01	02/14/2018	18:02
9	J1393-06	1	Solid	20.03	20	9.25	02/14/2018	18:03
10	J1402-02	1	Solid	20.03	20	8.35	02/14/2018	18:05
11	J1402-04	1	Solid	20.04	20	7.50	02/14/2018	18:06
12	J1402-06	1	Solid	20.02	20	7.33	02/14/2018	18:08
13	J1404-04	1	Solid	20.05	20	7.32	02/14/2018	18:10
14	J1550-01	1	Solid	20.03	20	9.35	02/14/2018	18:11
15	J1558-03	1	Solid	20.01	20	7.85	02/14/2018	18:13
16	CCV2	1	Water			12.01	02/14/2018	18:15
17	J1559-01	1	Solid	20.02	20	6.59	02/14/2018	18:16
18	J1559-02	1	Solid	20.03	20	7.95	02/14/2018	18:19
19	J1559-03	1	Solid	20.00	20	7.03	02/14/2018	18:20
20	J1562-07	1	Solid	20.02	20	6.59	02/14/2018	18:21

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water			2.01	02/14/2018	18:23

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY J1561

WorkList ID : 108556

Date : 2/14/2018 3:16:38 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/14/2018	Solid	J1393-02	Corrosivity	Cool 4 deg C	PSEG05	P53	HR-06-021418-A	02/14/2018	9045C
02/14/2018	Solid	J1393-04	Corrosivity	Cool 4 deg C	PSEG05	P53	HR-06-021418-B	02/14/2018	9045C
02/14/2018	Solid	J1393-06	Corrosivity	Cool 4 deg C	PSEG05	P53	HR-06-021418-C	02/14/2018	9045C
02/14/2018	Solid	J1402-02	Corrosivity	Cool 4 deg C	PSEG05	P61	PL-01-021418-A	02/14/2018	9045C
02/14/2018	Solid	J1402-04	Corrosivity	Cool 4 deg C	PSEG05	P61	PL-01-021418-B	02/14/2018	9045C
02/14/2018	Solid	J1402-06	Corrosivity	Cool 4 deg C	PSEG05	P61	PL-01-021418-C	02/14/2018	9045C
02/14/2018	Solid	J1404-04	Corrosivity	Cool 4 deg C	PSEG05		TR-04-021418	02/14/2018	9045C
02/13/2018	Solid	J1550-01	Corrosivity	Cool 4 deg C	SEAC01	P42	SOIL-CUTTINGS	02/13/2018	9045C
02/14/2018	Solid	J1558-03	Corrosivity	Cool 4 deg C	PSEG01	P52	RO-155	02/14/2018	9045C
02/13/2018	Solid	J1559-01	Corrosivity	Cool 4 deg C	CLEA11	P63	OWS-NEAR-FUEL-STATION	02/13/2018	9045C
02/13/2018	Solid	J1559-02	Corrosivity	Cool 4 deg C	CLEA11	P63	CHASSIS-WASH-SYSTEM	02/13/2018	9045C
02/13/2018	Solid	J1559-03	Corrosivity	Cool 4 deg C	CLEA11	P63	BUS-WASH-ZIPPER-DRAIN	02/13/2018	9045C
02/14/2018	Solid	J1562-07	Corrosivity		PSEG01	P52	T-10-STOCKPILE	02/14/2018	9045C

Date/Time 02-14-18 09:00 AM
 Received by: JP
 Relinquished by: CB

Date/Time 02-14-18 7:25 AM
 Received by: CB
 Relinquished by: JP