

7470A


LB 93297

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38003



A= 0.0000e+000

B= 2.6615e-004

C= 3.5965e-002

Rho= 0.9995093

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.019	-0.019	-205	0.000	-205				
0.2	0.200	0.237	0.037	756	0.0 %	756				
2.5	2.500	2.576	0.076	9543	0.0 %	9543				
5.0	5.000	4.978	-0.022	18569	0.0 %	18569				
7.5	7.500	7.277	-0.223	27208	0.0 %	27208				
10.0	10.000	10.150	0.150	38003	0.0 %	38003				

LB93297 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Feb 2018 11:18:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-205	-	0.0000	7470A	PPB	15 Feb 2018 11:37:04	S	Std
0.2 - 1	50.2	756	-	0.2000	7470A	PPB	15 Feb 2018 11:41:38	S	Std
2.5 - 1	52.5	9543	-	2.5000	7470A	PPB	15 Feb 2018 11:43:46	S	Std
5.0 - 1	55	18569	-	5.0000	7470A	PPB	15 Feb 2018 11:45:54	S	Std
7.5 - 1	57.5	27208	-	7.5000	7470A	PPB	15 Feb 2018 11:48:02	S	Std
10.0 - 1	510	38003	-	10.0000	7470A	PPB	15 Feb 2018 11:52:42	S	Std
ICV01 - 1	ICV01	15233	4.0902	-	7470A	PPB	15 Feb 2018 11:55:12	U	SMPL
ICB01 - 1	ICB01	-196	-0.0162	-	7470A	PPB	15 Feb 2018 11:57:18	U	SMPL
CCV50 - 1	CCV50	18743	5.0243	-	7470A	PPB	15 Feb 2018 11:59:26	U	SMPL
CCB50 - 1	CCB50	-195	-0.0159	-	7470A	PPB	15 Feb 2018 12:01:32	U	SMPL
CRA - 1	CRA	714	0.2260	-	7470A	PPB	15 Feb 2018 12:06:04	U	SMPL
HighStd - 1	HighStd	38804	10.3635	-	7470A	PPB	15 Feb 2018 12:08:10	U	SMPL
ChkStd - 1	ChkStd	25681	6.8709	-	7470A	PPB	15 Feb 2018 12:10:17	U	SMPL
PB106567BL - 1	PBW	-266	-0.0348	-	7470A	PPB	15 Feb 2018 12:12:26	U	SMPL
PB106567BS - 1	LCSW	13377	3.5962	-	7470A	PPB	15 Feb 2018 12:14:35	U	SMPL
PB106567TB - 1	PB106528TB	-202	-0.0178	-	7470A	PPB	15 Feb 2018 12:16:42	U	SMPL
J1530-41 - 1	WC-01	-141	-0.0016	-	7470A	PPB	15 Feb 2018 12:18:51	U	SMPL
J1530-41DUP - 1	WC-01DUP	-118	0.0046	-	7470A	PPB	15 Feb 2018 12:20:59	U	SMPL
J1530-41MS - 1	WC-01MS	13700	3.6822	-	7470A	PPB	15 Feb 2018 12:23:08	U	SMPL
J1530-41MSD - 1	WC-01MSD	14094	3.7870	-	7470A	PPB	15 Feb 2018 12:25:14	U	SMPL
CCV51 - 1	CCV51	18375	4.9264	-	7470A	PPB	15 Feb 2018 12:27:21	U	SMPL
CCB51 - 1	CCB51	-202	-0.0178	-	7470A	PPB	15 Feb 2018 12:29:28	U	SMPL
J1530-42 - 1	WC-02	-122	0.0035	-	7470A	PPB	15 Feb 2018 12:31:34	U	SMPL
J1530-43 - 1	WC-03	-101	0.0091	-	7470A	PPB	15 Feb 2018 12:33:41	U	SMPL
J1530-44 - 1	WC-04	-109	0.0070	-	7470A	PPB	15 Feb 2018 12:35:47	U	SMPL
J1541-04 - 1	SAMPLE-A-1-3	-86	0.0131	-	7470A	PPB	15 Feb 2018 12:37:54	U	SMPL
J1541-08 - 1	SAMPLE-B-1-3	-68	0.0179	-	7470A	PPB	15 Feb 2018 12:40:02	U	SMPL
J1541-12 - 1	SAMPLE-C-1-3	-59	0.0203	-	7470A	PPB	15 Feb 2018 12:42:09	U	SMPL
J1545-07 - 1	TEST-PIT-1	-108	0.0072	-	7470A	PPB	15 Feb 2018 12:44:17	U	SMPL
J1545-14 - 1	TEST-PIT-2	-91	0.0117	-	7470A	PPB	15 Feb 2018 12:46:26	U	SMPL
J1530-4LX5 - 1		-85	0.0133	-	7470A	PPB	15 Feb 2018 12:48:34	U	SMPL
J1530-41A - 1		13672	3.6747	-	7470A	PPB	15 Feb 2018 12:50:41	U	SMPL
CCV52 - 1	CCV52	18734	5.0219	-	7470A	PPB	15 Feb 2018 12:52:48	U	SMPL
CCB52 - 1	CCB52	-180	-0.0119	-	7470A	PPB	15 Feb 2018 12:54:55	U	SMPL