

7471B


LB 93 30 9

INSTRUMENT ID: CY1

Linear

 μ Abs.:

45366



A= 0.0000e+000

B= 2.2349e-004

C= -6.3600e-003

Rho= 0.9995217

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.006	-0.006	0	0.000	0				
0.2	0.200	0.219	0.019	1008	0.0 %	1008				
2.5	2.500	2.536	0.036	11376	0.0 %	11376				
5.0	5.000	5.055	0.055	22649	0.0 %	22649				
7.5	7.500	7.264	-0.236	32530	0.0 %	32530				
10	10.000	10.132	0.132	45366	0.0 %	45366				

LB93309 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 15 Feb 2018 15:21:59

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	0	-	0.0000	7471B	PPB	15 Feb 2018 15:24:37	S	Std
0.2 - 1	50.2	1008	-	0.2000	7471B	PPB	15 Feb 2018 15:26:54	S	Std
2.5 - 1	52.5	11376	-	2.5000	7471B	PPB	15 Feb 2018 15:29:13	S	Std
5.0 - 1	55	22649	-	5.0000	7471B	PPB	15 Feb 2018 15:31:30	S	Std
7.5 - 1	57.5	32530	-	7.5000	7471B	PPB	15 Feb 2018 15:33:47	S	Std
10 - 1	510	45366	-	10.0000	7471B	PPB	15 Feb 2018 15:36:03	S	Std
ICV02 - 1	ICV02	17306	3.8613	-	7471B	PPB	15 Feb 2018 15:38:48	U	SMPL
ICB02 - 1	ICB02	-136	-0.0368	-	7471B	PPB	15 Feb 2018 15:41:04	U	SMPL
CCV53 - 1	CCV53	22707	5.0684	-	7471B	PPB	15 Feb 2018 15:43:19	U	SMPL
CCB53 - 1	CCB53	-144	-0.0385	-	7471B	PPB	15 Feb 2018 15:45:35	U	SMPL
CRA - 1	CRA	973	0.2111	-	7471B	PPB	15 Feb 2018 15:47:50	U	SMPL
HighStd - 1	HighStd	45502	10.1627	-	7471B	PPB	15 Feb 2018 15:50:06	U	SMPL
ChkStd - 1	ChkStd	31209	6.9684	-	7471B	PPB	15 Feb 2018 15:52:22	U	SMPL
PB106574BL - 1	PBS	-281	-0.0692	-	7471B	PPB	15 Feb 2018 15:54:40	U	SMPL
PB106574BS - 1	LCSS	16424	3.6642	-	7471B	PPB	15 Feb 2018 15:56:58	U	SMPL
J1393-01 - 1	HR-06-021418-A	8828	1.9666	-	7471B	PPB	15 Feb 2018 15:59:15	U	SMPL
J1393-03 - 1	HR-06-021418-B	2283	0.5039	-	7471B	PPB	15 Feb 2018 16:01:32	U	SMPL
J1393-05 - 1	HR-06-021418-C	10687	2.3820	-	7471B	PPB	15 Feb 2018 16:03:49	U	SMPL
J1404-03 - 1	TR-04-021418	1626	0.3570	-	7471B	PPB	15 Feb 2018 16:06:07	U	SMPL
J1536-03 - 1	COMPOSITE-A-20180150	6184	1.3757	-	7471B	PPB	15 Feb 2018 16:08:23	U	SMPL
CCV54 - 1	CCV54	22584	5.0409	-	7471B	PPB	15 Feb 2018 16:10:39	U	SMPL
CCB54 - 1	CCB54	-146	-0.0390	-	7471B	PPB	15 Feb 2018 16:12:55	U	SMPL
J1536-04 - 1	COMPOSITE-B-20180151	5810	1.2921	-	7471B	PPB	15 Feb 2018 16:15:11	U	SMPL
J1541-01 - 1	SAMPLE-A-1-3	1266	0.2766	-	7471B	PPB	15 Feb 2018 16:17:28	U	SMPL
J1541-05 - 1	SAMPLE-B-1-3	1345	0.2942	-	7471B	PPB	15 Feb 2018 16:19:43	U	SMPL
J1541-09 - 1	SAMPLE-C-1-3	1206	0.2632	-	7471B	PPB	15 Feb 2018 16:22:00	U	SMPL
J1545-01 - 1	TEST-PIT-1	691	0.1481	-	7471B	PPB	15 Feb 2018 16:24:16	U	SMPL
J1545-08 - 1	TEST-PIT-2	747	0.1606	-	7471B	PPB	15 Feb 2018 16:26:33	U	SMPL
J1548-01 - 1	22A-2	28686	6.4046	-	7471B	PPB	15 Feb 2018 16:28:50	U	SMPL
J1548-01DUP - 1	22A-2DUP	28260	6.3094	-	7471B	PPB	15 Feb 2018 16:33:49	U	SMPL
J1548-01MS - 1	22A-2MS	44685	9.9802	-	7471B	PPB	15 Feb 2018 16:41:25	U	SMPL
J1548-01MSD - 1	22A-2MSD	44180	9.8673	-	7471B	PPB	15 Feb 2018 16:46:20	U	SMPL
CCV55 - 1	CCV55	22361	4.9910	-	7471B	PPB	15 Feb 2018 16:48:36	U	SMPL
CCB55 - 1	CCB55	-160	-0.0421	-	7471B	PPB	15 Feb 2018 16:50:54	U	SMPL
J1558-01 - 1	RO-155	492	0.1036	-	7471B	PPB	15 Feb 2018 16:53:11	U	SMPL
J1561-01 - 1	GL-WC-02	11628	2.5923	-	7471B	PPB	15 Feb 2018 16:55:27	U	SMPL
J1562-01 - 1	T-10-FOUNDATION	191	0.0363	-	7471B	PPB	15 Feb 2018 16:57:43	U	SMPL
J1562-03 - 1	T-10-STOCKPILE	7172	1.5965	-	7471B	PPB	15 Feb 2018 17:00:00	U	SMPL
J1548-01LX5 - 1		5449	1.2114	-	7471B	PPB	15 Feb 2018 17:02:16	U	SMPL
J1548-01A - 1		44892	10.0264	-	7471B	PPB	15 Feb 2018 17:04:33	U	SMPL
CCV56 - 1	CCV56	22636	5.0525	-	7471B	PPB	15 Feb 2018 17:12:28	U	SMPL
CCB56 - 1	CCB56	-156	-0.0412	-	7471B	PPB	15 Feb 2018 17:14:47	U	SMPL