



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By :

Run Number: LB93314

Temp.: 23.8 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/15/2018	18:10
2	CAL2	1	Water			6.99	02/15/2018	18:11
3	CAL3	1	Water			10.01	02/15/2018	18:12
4	ICV	1	Water			7.00	02/15/2018	18:14
5	CCV1	1	Water			2.01	02/15/2018	18:15
6	J1399-06	1	Solid	20.02	20	8.61	02/15/2018	18:16
7	J1399-06DUP	1	Solid	20.03	20	8.62	02/15/2018	18:19
8	J1399-08	1	Solid	20.04	20	8.46	02/15/2018	18:20
9	J1401-02	1	Solid	20.02	20	8.43	02/15/2018	18:22
10	J1401-04	1	Solid	20.03	20	8.39	02/15/2018	18:23
11	J1574-07	1	Solid	20.02	20	7.72	02/15/2018	18:25
12	J1575-06	1	Solid	20.05	20	5.45	02/15/2018	18:26
13	J1575-12	1	Solid	20.03	20	6.02	02/15/2018	18:28
14	CCV2	1	Water			12.02	02/15/2018	18:30

WORKLIST(Hardcopy Internal Chain)

20 93314

WorkList Name : CORROSIONITY J1575

WorkList ID : 108610

Date : 2/15/2018 3:24:11 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/15/2018	Solid	J1399-06	Corrosivity	Cool 4 deg C	PSEG05		OR-02-021518-A	02/15/2018	9045C
02/15/2018	Solid	J1399-08	Corrosivity	Cool 4 deg C	PSEG05		OR-02-021518-B	02/15/2018	9045C
02/15/2018	Solid	J1401-02	Corrosivity	Cool 4 deg C	PSEG05	P63	SA-06-021518-A	02/15/2018	9045C
02/15/2018	Solid	J1401-04	Corrosivity	Cool 4 deg C	PSEG05	P63	SA-06-021518-B	02/15/2018	9045C
02/15/2018	Solid	J1574-07	Corrosivity	Cool 4 deg C	PSEG02	P62	DPC-91733-OW-01	02/15/2018	9045C
02/15/2018	Solid	J1575-06	Corrosivity	Cool 4 deg C	PSEG01	P62	SAMPLE-A	02/15/2018	9045C
02/15/2018	Solid	J1575-12	Corrosivity	Cool 4 deg C	PSEG01	P62	SAMPLE-B	02/15/2018	9045C

Date/Time 02-15-18 3:30 PM
 Received by: JN
 Relinquished by: JN

Date/Time 02-15-18 6:35 PM
 Received by: CD
 Relinquished by: JN