



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Aparna

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB93321

Reagent/Standard	Lot/Log #
Residual chlorine std, 0.8PPM	WP62004
Residual chlorine std, 0PPM	WP62000
Residual chlorine std, 0.1PPM	WP62001
Residual chlorine std, 0.2PPM	WP62002
Residual chlorine std, and CCV , 0.4PPM	WP62003
Residual chlorine std, 1.6PPM	WP62005
Residual chlorine ICV-LCS, 0.4PPM	WP62006
Residual chlorine LOD, 0.05PPM	WP62007
Total Chlorine Powder Pillows	W1960

Intercept: -0.0003

Slope: 1.0102

Regression: 0.999956

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	02/16/2018	11:03
2	CAL2	0.1	1	0.000	0.100	0.100	02/16/2018	11:05
3	CAL3	0.2	1	0.000	0.210	0.210	02/16/2018	11:05
4	CAL4	0.4	1	0.000	0.400	0.400	02/16/2018	11:06
5	CAL5	0.8	1	0.000	0.800	0.800	02/16/2018	11:06
6	CAL6	1.6	1	0.000	1.620	1.620	02/16/2018	11:06



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Aparna

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB93321

Seq	Lab ID	Initial Weight	Final Vol	True Value	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.000	0.400	0.400	0.396	02/16/2018	11:07
2	ICB				1	0.000	0.000	0.000	0.000	02/16/2018	11:07
3	CCV1			0.4	1	0.000	0.440	0.440	0.436	02/16/2018	11:07
4	CCB1				1	0.000	0.000	0.000	0.000	02/16/2018	11:08
5	LB93321BL	50	50		1	0.000	0.000	0.000	0.000	02/16/2018	11:08
6	LB93321BS	50	50	0.4	1	0.000	0.400	0.400	0.396	02/16/2018	11:09
7	J1161-07	50	50		1	0.000	0.060	0.060	0.060	02/16/2018	11:09
8	J1161-09	50	50		1	0.000	0.050	0.050	0.050	02/16/2018	11:10
9	CCV2			0.4	1	0.000	0.420	0.420	0.416	02/16/2018	11:10
10	CCB2				1	0.000	0.000	0.000	0.000	02/16/2018	11:11