


7470A

LB93323
INSTRUMENT ID: CV1

Linear

 μ Abs.:

37690



A= 0.0000e+000

B= 2.6759e-004

C= -4.7981e-003

Rho= 0.9998219

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.018	-0.018	-50	0.000	-50				
0.2	0.200	0.209	0.009	800	0.0 %	800				
2.5	2.500	2.553	0.053	9557	0.0 %	9557				
5.0	5.000	5.015	0.015	18758	0.0 %	18758				
7.5	7.500	7.361	-0.139	27528	0.0 %	27528				
10.0	10.000	10.081	0.081	37690	0.0 %	37690				

LB93323 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Feb 2018 11:36:30

Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-50	-	0.00007470A	PPB	16 Feb 2018 11:44:59	S	Std
0.2 - 1	50.2	800	-	0.20007470A	PPB	16 Feb 2018 11:47:07	S	Std
2.5 - 1	52.5	9557	-	2.50007470A	PPB	16 Feb 2018 11:49:15	S	Std
5.0 - 1	55	18758	-	5.00007470A	PPB	16 Feb 2018 11:51:23	S	Std
7.5 - 1	57.5	27528	-	7.50007470A	PPB	16 Feb 2018 11:53:31	S	Std
10.0 - 1	510	37690	-	10.00007470A	PPB	16 Feb 2018 11:55:39	S	Std
ICV03 - 1	ICV03	14101	3.7684	-7470A	PPB	16 Feb 2018 11:58:17	U	SMPL
ICB03 - 1	ICB03	-176	-0.0519	-7470A	PPB	16 Feb 2018 12:00:23	U	SMPL
CCV57 - 1	CCV57	18996	5.0783	-7470A	PPB	16 Feb 2018 12:02:29	U	SMPL
CCB57 - 1	CCB57	-155	-0.0463	-7470A	PPB	16 Feb 2018 12:04:36	U	SMPL
CRA - 1	CRA	778	0.2034	-7470A	PPB	16 Feb 2018 12:06:42	U	SMPL
HighStd - 1	HighStd	37824	10.1164	-7470A	PPB	16 Feb 2018 12:08:49	U	SMPL
ChkStd - 1	ChkStd	25822	6.9048	-7470A	PPB	16 Feb 2018 12:10:56	U	SMPL
PB106602BL - 1	PBW	-258	-0.0738	-7470A	PPB	16 Feb 2018 12:13:04	U	SMPL
PB106602BS - 1	LCSW	13109	3.5030	-7470A	PPB	16 Feb 2018 12:15:13	U	SMPL
PB106602TB - 1	PB106568TB	-162	-0.0481	-7470A	PPB	16 Feb 2018 12:17:20	U	SMPL
J1393-02 - 1	HR-06-021418-A	-141	-0.0425	-7470A	PPB	16 Feb 2018 12:19:28	U	SMPL
J1393-04 - 1	HR-06-021418-B	-72	-0.0241	-7470A	PPB	16 Feb 2018 12:21:36	U	SMPL
J1393-06 - 1	HR-06-021418-C	-104	-0.0326	-7470A	PPB	16 Feb 2018 12:23:45	U	SMPL
J1402-02 - 1	PL-01-021418-A	-78	-0.0257	-7470A	PPB	16 Feb 2018 12:25:51	U	SMPL
CCV58 - 1	CCV58	18960	5.0686	-7470A	PPB	16 Feb 2018 12:27:58	U	SMPL
CCB58 - 1	CCB58	-148	-0.0444	-7470A	PPB	16 Feb 2018 12:30:05	U	SMPL
J1402-04 - 1	PL-01-021418-B	-83	-0.0270	-7470A	PPB	16 Feb 2018 12:32:12	U	SMPL
J1402-06 - 1	PL-01-021418-C	-85	-0.0275	-7470A	PPB	16 Feb 2018 12:34:18	U	SMPL
J1404-04 - 1	TR-04-021418	-108	-0.0337	-7470A	PPB	16 Feb 2018 12:36:25	U	SMPL
J1550-01 - 1	SOIL-CUTTINGS	-31	-0.0131	-7470A	PPB	16 Feb 2018 12:38:32	U	SMPL
J1558-03 - 1	RO-155	-95	-0.0302	-7470A	PPB	16 Feb 2018 12:40:39	U	SMPL
J1558-03DUP - 1	RO-155DUP	-85	-0.0275	-7470A	PPB	16 Feb 2018 12:42:47	U	SMPL
J1558-03MS - 1	RO-155MS	13102	3.5011	-7470A	PPB	16 Feb 2018 12:44:55	U	SMPL
J1558-03MSD - 1	RO-155MSD	13070	3.4925	-7470A	PPB	16 Feb 2018 12:49:34	U	SMPL
J1559-01 - 1	OWS-NEAR-FUEL-STATION	-97	-0.0308	-7470A	PPB	16 Feb 2018 12:51:42	U	SMPL
J1559-02 - 1	CHASSIS-WASH-SYSTEM	-110	-0.0342	-7470A	PPB	16 Feb 2018 12:53:49	U	SMPL
CCV59 - 1	CCV59	18870	5.0445	-7470A	PPB	16 Feb 2018 12:55:56	U	SMPL
CCB59 - 1	CCB59	-198	-0.0578	-7470A	PPB	16 Feb 2018 12:58:03	U	SMPL
J1559-03 - 1	BUS-WASH-ZIPPER-DRAIN	-77	-0.0254	-7470A	PPB	16 Feb 2018 13:00:10	U	SMPL
J1560-01 - 1	15225-7	-99	-0.0313	-7470A	PPB	16 Feb 2018 13:02:17	U	SMPL
J1560-05 - 1	15202-3	-72	-0.0241	-7470A	PPB	16 Feb 2018 13:04:24	U	SMPL
J1560-09 - 1	15228-9	543	0.1405	-7470A	PPB	16 Feb 2018 13:06:31	U	SMPL
J1561-02 - 1	GL-WC-02	-89	-0.0286	-7470A	PPB	16 Feb 2018 13:08:38	U	SMPL
J1562-02 - 1	T-10-FOUNDATION	-90	-0.0289	-7470A	PPB	16 Feb 2018 13:10:46	U	SMPL
J1562-07 - 1	T-10-STOCKPILE	-46	-0.0171	-7470A	PPB	16 Feb 2018 13:12:54	U	SMPL
J1565-02 - 1	15340	-43	-0.0163	-7470A	PPB	16 Feb 2018 13:15:02	U	SMPL
J1565-11 - 1	19558	-98	-0.0310	-7470A	PPB	16 Feb 2018 13:17:11	U	SMPL
J1566-01 - 1	1451	-89	-0.0286	-7470A	PPB	16 Feb 2018 13:19:17	U	SMPL
CCV60 - 1	CCV60	18235	4.8746	-7470A	PPB	16 Feb 2018 13:21:24	U	SMPL
CCB60 - 1	CCB60	-195	-0.0570	-7470A	PPB	16 Feb 2018 13:23:31	U	SMPL
PB106603BL - 1	PBW	-115	-0.0356	-7470A	PPB	16 Feb 2018 13:25:39	U	SMPL
PB106603BS - 1	LCSW	12838	3.4305	-7470A	PPB	16 Feb 2018 13:27:46	U	SMPL
PB106603TB - 1	PB106572TB	-196	-0.0572	-7470A	PPB	16 Feb 2018 13:29:54	U	SMPL
J1565-01 - 1	18273	-89	-0.0286	-7470A	PPB	16 Feb 2018 13:32:01	U	SMPL
J1565-01DUP - 1	18273DUP	-84	-0.0273	-7470A	PPB	16 Feb 2018 13:34:08	U	SMPL
J1565-01MS - 1	18273MS	12050	3.2196	-7470A	PPB	16 Feb 2018 13:36:16	U	SMPL
J1565-01MSD - 1	18273MSD	12196	3.2587	-7470A	PPB	16 Feb 2018 13:38:24	U	SMPL
J1566-03 - 1	1452-56	-172	-0.0508	-7470A	PPB	16 Feb 2018 13:40:32	U	SMPL
J1566-05 - 1	1457-61	-80	-0.0262	-7470A	PPB	16 Feb 2018 13:42:40	U	SMPL
PB106604BL - 1	PBW	-52	-0.0187	-7470A	PPB	16 Feb 2018 13:44:47	U	SMPL
CCV61 - 1	CCV61	18567	4.9635	-7470A	PPB	16 Feb 2018 13:46:54	U	SMPL
CCB61 - 1	CCB61	-166	-0.0492	-7470A	PPB	16 Feb 2018 13:49:01	U	SMPL
PB106604BS - 1	LCSW	13521	3.6132	-7470A	PPB	16 Feb 2018 13:51:09	U	SMPL
J1343-01 - 1	MH-6	-283	-0.0805	-7470A	PPB	16 Feb 2018 13:55:56	U	SMPL
J1364-01 - 1	MW-03	3833	1.0209	-7470A	PPB	16 Feb 2018 13:58:03	U	SMPL
J1364-02 - 1	MW-04	1319	0.3481	-7470A	PPB	16 Feb 2018 14:00:10	U	SMPL
J1364-03 - 1	MW-05	10642	2.8428	-7470A	PPB	16 Feb 2018 14:02:17	U	SMPL
J1364-04 - 1	MW-07	627	0.1630	-7470A	PPB	16 Feb 2018 14:04:24	U	SMPL
J1364-05 - 1	MW-05-DUP	3454	0.9194	-7470A	PPB	16 Feb 2018 14:06:32	U	SMPL
J1364-06 - 1	EQUIPMENT-BLANK	-329	-0.0928	-7470A	PPB	16 Feb 2018 14:08:41	U	SMPL
J1364-08 - 1	MW-03	-270	-0.0770	-7470A	PPB	16 Feb 2018 14:10:49	U	SMPL
J1364-09 - 1	MW-04	-277	-0.0789	-7470A	PPB	16 Feb 2018 14:12:56	U	SMPL
CCV62 - 1	CCV62	18426	4.9257	-7470A	PPB	16 Feb 2018 14:15:04	U	SMPL
CCB62 - 1	CCB62	-194	-0.0567	-7470A	PPB	16 Feb 2018 14:17:11	U	SMPL
J1364-10 - 1	MW-05	-283	-0.0805	-7470A	PPB	16 Feb 2018 14:19:18	U	SMPL
J1364-11 - 1	MW-07	-276	-0.0787	-7470A	PPB	16 Feb 2018 14:21:25	U	SMPL
J1364-12 - 1	MW-05-DUP	-267	-0.0762	-7470A	PPB	16 Feb 2018 14:23:33	U	SMPL
J1364-13 - 1	EQUIPMENT-BLANK	-311	-0.0880	-7470A	PPB	16 Feb 2018 14:25:40	U	SMPL
J1365-02 - 1	GROUNDWATER	-300	-0.0851	-7470A	PPB	16 Feb 2018 14:27:47	U	SMPL
J1536-01 - 1	FIELD-BLANK-20180148	-42	-0.0160	-7470A	PPB	16 Feb 2018 14:29:55	U	SMPL
J1536-02 - 1	SITE-WATER-20180149	-3	-0.0056	-7470A	PPB	16 Feb 2018 14:32:02	U	SMPL
J1563-05 - 1	20180141	-330	-0.0931	-7470A	PPB	16 Feb 2018 14:34:10	U	SMPL
J1563-05DUP - 1	20180141DUP	-311	-0.0880	-7470A	PPB	16 Feb 2018 14:36:19	U	SMPL
J1563-05MS - 1	20180141MS	12409	3.3157	-7470A	PPB	16 Feb 2018 14:38:27	U	SMPL
CCV63 - 1	CCV63	18367	4.9099	-7470A	PPB	16 Feb 2018 14:40:34	U	SMPL

LB93323 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Feb 2018 11:36:30

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB63 - 1	CCB63	-188	-0.0551		-7470A	PPB	16 Feb 2018 14:42:42	U	SMPL
J1563-05MSD - 1	20180141MSD	12317	3.2911		-7470A	PPB	16 Feb 2018 14:44:50	U	SMPL
J1558-03LX5 - 1		-177	-0.0522		-7470A	PPB	16 Feb 2018 14:46:58	U	SMPL
J1558-03A - 1		13606	3.6360		-7470A	PPB	16 Feb 2018 14:49:05	U	SMPL
J1565-01LX5 - 1		-199	-0.0580		-7470A	PPB	16 Feb 2018 14:51:13	U	SMPL
J1565-01A - 1		12728	3.4010		-7470A	PPB	16 Feb 2018 14:53:21	U	SMPL
J1563-05LX5 - 1		-259	-0.0741		-7470A	PPB	16 Feb 2018 14:55:29	U	SMPL
J1563-05A - 1		13245	3.5394		-7470A	PPB	16 Feb 2018 14:57:36	U	SMPL
CCV64 - 1	CCV64	18494	4.9439		-7470A	PPB	16 Feb 2018 14:59:44	U	SMPL
CCB64 - 1	CCB64	-189	-0.0554		-7470A	PPB	16 Feb 2018 15:01:53	U	SMPL