

LB93330

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AS

2/16/2018 12:19

Test: Phenolics-mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.026	0.0	0.115	
ICB1	0.002	0.0	0.009	
CCV1	1.071	0.0	0.120	
CCB1	0.002	0.0	0.009	
PB106599BL	-0.003	0.0	0.009	
PB106599BS	1.008	0.0	0.113	
J1161-10	0.007	0.0	0.010	
J1161-12	0.008	0.0	0.010	
CCV2	1.057	0.0	0.118	
CCB2	0.000	0.0	0.009	

N	10
Mean	0.418
SD	0.5360
CV%	128.31

AS

Aquakem v. 7.2AQ1

Results from time period:

Fri Feb 16 12:17:15 2018

Fri Feb 16 12:17:24 2018

Sample Id	Sam/Ctr/c/	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0518	mg/l	2/16/2018 10:53:56	
0.1PPM	A	Phenolics-	P	0.1029	mg/l	2/16/2018 10:53:57	
0.25PPM	A	Phenolics-	P	0.2501	mg/l	2/16/2018 10:53:58	
0.50PPM	A	Phenolics-	P	0.4974	mg/l	2/16/2018 10:53:59	
1.0PPM	A	Phenolics-	P	0.9946	mg/l	2/16/2018 10:54:00	
2.0PPM	A	Phenolics-	P	2.0031	mg/l	2/16/2018 10:54:01	
ICV1	S	Phenolics-	P	1.026	mg/l	2/16/2018 12:17:15	
ICB1	S	Phenolics-	P	0.0018	mg/l	2/16/2018 12:17:16	
CCV1	S	Phenolics-	P	1.0706	mg/l	2/16/2018 12:17:17	
CCB1	S	Phenolics-	P	0.0015	mg/l	2/16/2018 12:17:18	
PB106599BL	S	Phenolics-	P	-0.0027	mg/l	2/16/2018 12:17:19	
PB106599BS	S	Phenolics-	P	1.0076	mg/l	2/16/2018 12:17:20	
J1161-10	S	Phenolics-	P	0.0073	mg/l	2/16/2018 12:17:21	
J1161-12	S	Phenolics-	P	0.0082	mg/l	2/16/2018 12:17:22	
CCV2	S	Phenolics-	P	1.0567	mg/l	2/16/2018 12:17:23	
CCB2	S	Phenolics-	P	0.0003	mg/l	2/16/2018 12:17:24	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

2/16/2018 11:02


Test Phenolics-

Accepted 2/16/2018 11:02

Factor 9.671
Bias 0.009

Coeff. of det. 0.999980

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.014	0.0518	0.0500	
2	Phenol-2	0.020	0.1029	0.1000	
3	Phenol-2	0.035	0.2501	0.2500	
4	Phenol-2	0.060	0.4974	0.5000	
5	Phenol-2	0.112	0.9946	1.0000	
6	Phenol-2	0.216	2.0031	2.0000	