



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB93341

Temp.: 23.3 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	02/16/2018	14:30
2	CAL2	1	Water			7.00	02/16/2018	14:32
3	CAL3	1	Water			10.01	02/16/2018	14:34
4	ICV	1	Water			7.00	02/16/2018	14:36
5	CCV1	1	Water			2.00	02/16/2018	14:38
6	J1601-03	1	Water			6.02	02/16/2018	14:40
7	J1601-03DUP	1	Water			6.03	02/16/2018	14:42
8	CCV2	1	Water			12.00	02/16/2018	14:45

WORKLIST(Hardcopy Internal Chain)

LB 93341

WorkList Name : corrosivity021618

WorkList ID : 108656

Date : 2/16/2018 2:18:39 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/16/2018	Water	J1601-03 B	Corrosivity	Cool 4 deg C	PSEG01	P63	DRUM-1	02/16/2018	9040C

Date/Time

02/16/18 2:25 PM

Received by:

SP

Relinquished by:

SP

Date/Time

02/16/18 2:55 PM

Received by:

SP

Relinquished by:

SP