



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB93348

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/16/2018	16:30
2	CAL2	1	Water			7.01	02/16/2018	16:31
3	CAL3	1	Water			10.03	02/16/2018	16:33
4	ICV	1	Water			7.00	02/16/2018	16:35
5	CCV1	1	Water			2.01	02/16/2018	16:36
6	J1392-02	1	Solid	20.02	20	8.88	02/16/2018	16:37
7	J1604-03	1	Solid	20.04	20	7.35	02/16/2018	16:38
8	J1604-06	1	Solid	20.02	20	8.50	02/16/2018	16:40
9	J1604-09	1	Solid	20.01	20	7.66	02/16/2018	16:41
10	J1604-12	1	Solid	20.05	20	6.89	02/16/2018	16:42
11	J1604-12DUP	1	Solid	20.03	20	6.90	02/16/2018	16:44
12	CCV2	1	Water			12.02	02/16/2018	16:45

LB 93348

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY J1604 WorkList ID : 108661 Date : 2/16/2018 2:49:51 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/16/2018	Solid	J1392-02	Corrosivity	Cool 4 deg C	PSEG05	P63	BUJ-03-021618	02/16/2018	9045C
02/16/2018	Solid	J1604-03	Corrosivity	Cool 4 deg C	PSEG01	P63	TP-1	02/16/2018	9045C
02/16/2018	Solid	J1604-06	Corrosivity	Cool 4 deg C	PSEG01	P63	TP-2	02/16/2018	9045C
02/16/2018	Solid	J1604-09	Corrosivity	Cool 4 deg C	PSEG01	P63	TP-3	02/16/2018	9045C
02/16/2018	Solid	J1604-12	Corrosivity	Cool 4 deg C	PSEG01	P63	TP-4	02/16/2018	9045C

Date/Time 02-16-18 3:00 PM
Received by: JD
Relinquished by: JD

Date/Time 02-16-18 5:35 PM
Received by: CS
Relinquished by: JD