


7471B

LB93354
INSTRUMENT ID: CV1

Linear

 μ Abs.:

45601



A= 0.0000e+000

B= 2.2165e-004

C= -2.0829e-002

Rho= 0.9997621

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.022	-0.022	-4	0.000	-4				
0.2	0.200	0.197	-0.003	985	0.0 %	985				
2.5	2.500	2.582	0.082	11742	0.0 %	11742				
5.0	5.000	5.011	0.011	22702	0.0 %	22702				
7.5	7.500	7.345	-0.155	33231	0.0 %	33231				
10	10.000	10.087	0.087	45601	0.0 %	45601				

LB93354 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Feb 2018 09:27:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-4	-	0.0000	7471B	PPB	19 Feb 2018 10:01:20	S	Std
0.2 - 1	50.2	985	-	0.2000	7471B	PPB	19 Feb 2018 10:03:38	S	Std
2.5 - 1	52.5	11742	-	2.5000	7471B	PPB	19 Feb 2018 10:05:56	S	Std
5.0 - 1	55	22702	-	5.0000	7471B	PPB	19 Feb 2018 10:08:13	S	Std
7.5 - 1	57.5	33231	-	7.5000	7471B	PPB	19 Feb 2018 10:10:30	S	Std
10 - 1	510	45601	-	10.0000	7471B	PPB	19 Feb 2018 10:12:46	S	Std
ICV05 - 1	ICV05	18441	4.0666	-7471B	PPB	PPB	19 Feb 2018 10:15:33	U	SMPL
ICB05 - 1	ICB05	-151	-0.0543	-7471B	PPB	PPB	19 Feb 2018 10:17:48	U	SMPL
CCV68 - 1	CCV68	22816	5.0363	-7471B	PPB	PPB	19 Feb 2018 10:20:03	U	SMPL
CCB68 - 1	CCB68	-115	-0.0463	-7471B	PPB	PPB	19 Feb 2018 10:22:19	U	SMPL
CRA - 1	CRA	986	0.1977	-7471B	PPB	PPB	19 Feb 2018 10:24:36	U	SMPL
HighStd - 1	HighStd	45618	10.0904	-7471B	PPB	PPB	19 Feb 2018 10:26:51	U	SMPL
ChkStd - 1	ChkStd	32935	7.2792	-7471B	PPB	PPB	19 Feb 2018 10:29:08	U	SMPL
PB106625BL - 1	PBS	-261	-0.0787	-7471B	PPB	PPB	19 Feb 2018 10:31:26	U	SMPL
PB106625BS - 1	LCSS	16091	3.5457	-7471B	PPB	PPB	19 Feb 2018 10:33:44	U	SMPL
J1399-05 - 1	OR-02-021518-A	1601	0.3340	-7471B	PPB	PPB	19 Feb 2018 10:36:01	U	SMPL
J1399-07 - 1	OR-02-021518-B	3345	0.7206	-7471B	PPB	PPB	19 Feb 2018 10:38:18	U	SMPL
J1401-01 - 1	SA-06-021518-A	5252	1.1433	-7471B	PPB	PPB	19 Feb 2018 10:40:36	U	SMPL
J1401-03 - 1	SA-06-021518-B	9223	2.0234	-7471B	PPB	PPB	19 Feb 2018 10:42:54	U	SMPL
J1402-01 - 1	PL-01-021418-A	443	0.0774	-7471B	PPB	PPB	19 Feb 2018 10:45:10	U	SMPL
CCV69 - 1	CCV69	22585	4.9851	-7471B	PPB	PPB	19 Feb 2018 10:47:25	U	SMPL
CCB69 - 1	CCB69	-174	-0.0594	-7471B	PPB	PPB	19 Feb 2018 10:49:41	U	SMPL
J1402-03 - 1	PL-01-021418-B	966	0.1933	-7471B	PPB	PPB	19 Feb 2018 10:51:58	U	SMPL
J1402-03DUP - 1	PL-01-021418-BDUP	833	0.1638	-7471B	PPB	PPB	19 Feb 2018 10:54:13	U	SMPL
J1402-03MS - 1	PL-01-021418-BMS	17720	3.9068	-7471B	PPB	PPB	19 Feb 2018 10:56:29	U	SMPL
J1402-03MSD - 1	PL-01-021418-BMSD	17774	3.9188	-7471B	PPB	PPB	19 Feb 2018 10:58:46	U	SMPL
J1402-05 - 1	PL-01-021418-C	491	0.0880	-7471B	PPB	PPB	19 Feb 2018 11:01:02	U	SMPL
J1575-01 - 1	SAMPLE-A	1230	0.2518	-7471B	PPB	PPB	19 Feb 2018 11:03:19	U	SMPL
J1575-07 - 1	SAMPLE-B	766	0.1490	-7471B	PPB	PPB	19 Feb 2018 11:05:37	U	SMPL
J1584-01 - 1	FK-GF-02-012-A	2600	0.5555	-7471B	PPB	PPB	19 Feb 2018 11:07:54	U	SMPL
J1584-03 - 1	FK-GF-02-012-B	2351	0.5003	-7471B	PPB	PPB	19 Feb 2018 11:10:12	U	SMPL
J1603-01 - 1	ORI-TS	1895	0.3992	-7471B	PPB	PPB	19 Feb 2018 11:12:28	U	SMPL
CCV70 - 1	CCV70	22853	5.0445	-7471B	PPB	PPB	19 Feb 2018 11:14:44	U	SMPL
CCB70 - 1	CCB70	-181	-0.0609	-7471B	PPB	PPB	19 Feb 2018 11:17:01	U	SMPL
J1604-01 - 1	TP-1	1478	0.3068	-7471B	PPB	PPB	19 Feb 2018 11:19:17	U	SMPL
J1604-04 - 1	TP-2	1197	0.2445	-7471B	PPB	PPB	19 Feb 2018 11:21:33	U	SMPL
J1604-07 - 1	TP-3	975	0.1953	-7471B	PPB	PPB	19 Feb 2018 11:23:49	U	SMPL
J1604-10 - 1	TP-4	923	0.1838	-7471B	PPB	PPB	19 Feb 2018 11:26:05	U	SMPL
J1402-03LX5 - 1		98	0.0009	-7471B	PPB	PPB	19 Feb 2018 11:28:22	U	SMPL
J1402-03A - 1		17789	3.9221	-7471B	PPB	PPB	19 Feb 2018 11:30:39	U	SMPL
CCV71 - 1	CCV71	22454	4.9561	-7471B	PPB	PPB	19 Feb 2018 11:32:56	U	SMPL
CCB71 - 1	CCB71	-143	-0.0525	-7471B	PPB	PPB	19 Feb 2018 11:35:13	U	SMPL