



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB93358
Analysis Date: 02/19/2018
BalanceID: WC SC-3
OvenID: EX Oven-2

ANALYST: Sweta
REVIEWED BY: apatel
Extraction Date: 02/19/2018
Extraction IN Time: 10:40
Extraction OUT Time: 11:40

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB93358BL	LB93358BL	WATER	2	1000	100	2.9874	2.9874	3.02	2.9874	2.9874	0	0
2	LB93358BS	LB93358BS	WATER	2	1000	100	2.9956	2.9956	3.01	3.0138	3.0138	0.0182	18.2
3	LB93358BSD	LB93358BSD	WATER	2	1000	100	2.9945	2.9945	3.00	3.0126	3.0126	0.0181	18.1
4	J1601-02	DRUM-1	WATER	2	1000	100	3.0795	3.0795	3.04	3.0799	3.0799	0.0004	0.4



QC Batch# LB93358 Test: TPH Analysis Date: 02/19/2018

Chemicals Used:

Chemical Name	Chemical Lot #
Hexane	W2317
pH Paper 0-14	W2248
Sodium Sulfate	EP1822
1:1 HCL	WP59249
Silica Gel	W2199
Sand	N/A

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP58006
LCSWD	10.00 ML	WP58005
MS/MSD	N/A	N/A

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance: 0.0021 In OVEN TEMP : 70 Dessicator Time In : 12:45
1.0000 gram Balance: 1.0005 In Time: 12:00
Bal Check Time: 10:55

After Analysis

0.0020 gram Balance: 0.0018 Out OVEN TEMP: 70 Dessicator Time Out: 13:55
1.0000 gram Balance: 1.0003 Out Time: 13:30
Bal Check Time: 14:20

WORKLIST(Hardcopy Internal Chain)

LB 93358

Worklist Name : TPHJ1601

Worklist ID : 108692

Date : 2/19/2018 10:29:57 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/21/2018	Water	J1601-02	TPH	Conc H2SO4 to	PSEG01	P63	DRUM-1	02/16/2018	1664A

Date/Time

02/19/18 10:35 AM

Received by:

SP

Relinquished by:

CP

Date/Time

02/19/18 2:25 PM

Received by:

CP

Relinquished by:

SP