



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB93360

Temp.: 24.3 °C

SLOPE.: 99.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/19/2018	11:20
2	CAL2	1	Water			7.00	02/19/2018	11:22
3	CAL3	1	Water			10.01	02/19/2018	11:24
4	ICV	1	Water			7.00	02/19/2018	11:26
5	CCV1	1	Water			2.00	02/19/2018	11:28
6	J1611-01	1	Water			6.80	02/19/2018	11:30
7	J1611-01DUP	1	Water			6.81	02/19/2018	11:32
8	J1611-02	1	Water			7.30	02/19/2018	11:34
9	CCV2	1	Water			12.00	02/19/2018	11:36

WORKLIST(Hardcopy Internal Chain)

LB 93366

Worklist Name : pH1611

Worklist ID : 108697

Date : 2/19/2018 11:10:59 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/18/2018	Water	J1611-01-D	pH	Cool 4 deg C	HOMEO1	P51	APT	02/18/2018	SM 4500-PH B
02/18/2018	Water	J1611-02-D	pH	Cool 4 deg C	HOMEO1	P51	BASEMENT	02/18/2018	SM 4500-PH B

Date/Time

02/19/18 11:15 AM

Received by:

SP

Relinquished by:

SP

Date/Time

02/19/18 11:40 AM

Received by:

SP

Relinquished by:

SP