



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB93361

Reagent/Standard	Lot/Log #
Residual chlorine std, 0.8PPM	WP62051
Residual chlorine std, 0PPM	WP62047
Residual chlorine std, 0.1PPM	WP62048
Residual chlorine std, 0.2PPM	WP62049
Residual chlorine std, and CCV , 0.4PPM	WP62050
Residual chlorine std, 1.6PPM	WP62052
Residual chlorine ICV-LCS, 0.4PPM	WP62053
Residual chlorine LOD, 0.05PPM	WP62054
Total Chlorine Powder Pillows	W1960

Intercept: -0.0071

Slope: 1.0299

Regression: 0.999880

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	02/19/2018	12:10
2	CAL2	0.1	1	0.000	0.100	0.100	02/19/2018	12:10
3	CAL3	0.2	1	0.000	0.200	0.200	02/19/2018	12:11
4	CAL4	0.4	1	0.000	0.400	0.400	02/19/2018	12:11
5	CAL5	0.8	1	0.000	0.800	0.800	02/19/2018	12:12
6	CAL6	1.6	1	0.000	1.650	1.650	02/19/2018	12:12



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB93361

Seq	Lab ID	Initial Weight	Final Vol	True Value	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.000	0.390	0.390	0.386	02/19/2018	12:13
2	ICB				1	0.000	0.000	0.000	0.007	02/19/2018	12:13
3	CCV1			0.4	1	0.000	0.400	0.400	0.395	02/19/2018	12:14
4	CCB1				1	0.000	0.000	0.000	0.007	02/19/2018	12:14
5	LB93361BL	50	50		1	0.000	0.000	0.000	0.007	02/19/2018	12:14
6	LB93361BS	50	50	0.4	1	0.000	0.400	0.400	0.395	02/19/2018	12:15
7	J1161-07	50	50		1	0.000	0.050	0.050	0.055	02/19/2018	12:16
8	J1161-09	50	50		1	0.000	0.050	0.050	0.055	02/19/2018	12:16
9	J1611-01	50	50		1	0.000	0.000	0.000	0.007	02/19/2018	12:17
10	J1611-02	50	50		1	0.000	0.860	0.860	0.842	02/19/2018	12:17
11	J1611-02DUP	50	50		1	0.000	0.860	0.860	0.842	02/19/2018	12:17
12	CCV2			0.4	1	0.000	0.410	0.410	0.405	02/19/2018	12:18
13	CCB2				1	0.000	0.000	0.000	0.007	02/19/2018	12:18

WORKLIST(Hardcopy Internal Chain)

WorkList Name : RCL021918

WorkList ID : 108703

Date : 2/19/2018 11:27:41 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/18/2018	Water	J1611-01	Residual Chlorine	Cool 4 deg C	HOME01	P51	APT	02/18/2018	SM4500 Cl G
02/18/2018	Water	J1611-02	Residual Chlorine	Cool 4 deg C	HOME01	P51	BASEMENT	02/18/2018	SM4500 Cl G

Date/Time 2-19-18 11:40
 Received by: AB
 Relinquished by: AB

Date/Time 2-19-18 12:30
 Received by: AB
 Relinquished by: AB