

LB93366

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

2/19/2018 12:01

Test: Phenolics-mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.996	0.0	0.115	
ICB1	-0.009	0.0	0.009	
CCV1	1.022	0.0	0.118	
CCB1	-0.005	0.0	0.010	
PB106638BL	-0.005	0.0	0.010	
PB106638BS	0.995	0.0	0.115	
J1591-01	0.003	0.0	0.011	
J1591-01DUP	-0.002	0.0	0.010	
J1591-01MS	0.706	0.0	0.085	
J1591-01MSD	0.707	0.0	0.085	
J1591-02	-0.004	0.0	0.010	
CCV2	1.035	0.0	0.119	
CCB2	-0.003	0.0	0.010	

N	13
Mean	0.418
SD	0.4850
CV%	116.04

AS

Aquakem v. 7.2AQ1

Results from time period:

Mon Feb 19 11:17:38 2018

Mon Feb 19 11:59:25 2018

Sample Id	Sam/Ctr/c/	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0575	mg/l	2/19/2018 11:17:38	
0.1PPM	A	Phenolics-	P	0.0941	mg/l	2/19/2018 11:17:39	
0.25PPM	A	Phenolics-	P	0.2471	mg/l	2/19/2018 11:17:40	
0.50PPM	A	Phenolics-	P	0.5035	mg/l	2/19/2018 11:17:41	
1.0PPM	A	Phenolics-	P	0.9964	mg/l	2/19/2018 11:17:42	
2.0PPM	A	Phenolics-	P	2.0014	mg/l	2/19/2018 11:17:43	
ICV1	S	Phenolics-	P	0.996	mg/l	2/19/2018 11:40:52	
ICB1	S	Phenolics-	P	-0.0091	mg/l	2/19/2018 11:40:53	
CCV1	S	Phenolics-	P	1.0215	mg/l	2/19/2018 11:40:54	
CCB1	S	Phenolics-	P	-0.0052	mg/l	2/19/2018 11:40:55	
PB106638BL	S	Phenolics-	P	-0.0052	mg/l	2/19/2018 11:40:56	
PB106638BS	S	Phenolics-	P	0.9949	mg/l	2/19/2018 11:40:57	
J1591-01	S	Phenolics-	P	0.003	mg/l	2/19/2018 11:59:18	
J1591-01DUP	S	Phenolics-	P	-0.0019	mg/l	2/19/2018 11:59:19	
J1591-01MS	S	Phenolics-	P	0.7055	mg/l	2/19/2018 11:59:20	
J1591-01MSD	S	Phenolics-	P	0.7067	mg/l	2/19/2018 11:59:21	
J1591-02	S	Phenolics-	P	-0.0042	mg/l	2/19/2018 11:59:22	
CCV2	S	Phenolics-	P	1.0349	mg/l	2/19/2018 11:59:23	
CCB2	S	Phenolics-	P	-0.0033	mg/l	2/19/2018 11:59:24	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

2/19/2018 11:20


Test Phenolics-

Accepted 2/19/2018 11:20

Factor 9.487
Bias 0.010

Coeff. of det. 0.999954

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.016	0.0575	0.0500	
2	Phenol-2	0.020	0.0941	0.1000	
3	Phenol-2	0.036	0.2471	0.2500	
4	Phenol-2	0.063	0.5035	0.5000	
5	Phenol-2	0.115	0.9964	1.0000	
6	Phenol-2	0.221	2.0014	2.0000	