



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: AparnaDate: 02/20/2018DATE & TIME IN: 02/19/2018 14:15OVEN TEMP: 105 °CRun Number: LB93370DATE & TIME OUT: 02/20/2018 11:30OVEN TEMP: 105 °CBalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB93370BL	LB93370BL	74.5332	74.5332	100	74.5332	74.5332	74.5332	0	0
2	J1591-01	MH-27	79.9834	79.9834	100	80.1018	80.1018	80.1018	0.1184	1184
3	J1591-01DUP	MH-27DUP	83.3707	83.3707	100	83.4884	83.4884	83.4884	0.1177	1177
4	J1591-02	MH-28	63.8298	63.8298	100	64.0588	64.0588	64.0588	0.229	2290

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS021918

WorkList ID : 108706

Date : 2/19/2018 11:51:14 AM

2193370

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/23/2018	Water	J1591-01	TS	Cool 4 deg C	DIST01	P41	MH-27	02/16/2018	SM2540 B
02/23/2018	Water	J1591-02	TS	Cool 4 deg C	DIST01	P41	MH-28	02/16/2018	SM2540 B

Date/Time 02/19/18 14:06

Received by: AS

Relinquished by:

Date/Time 02/19/18 14:45

Received by: AS

Relinquished by: AS