

7470A


LB 93379

INSTRUMENT ID: CV2

Linear

 μ Abs.:

37817



A= 0.0000e+000

B= 2.6807e-004

C= 5.9487e-004

Rho= 0.9996193

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	10	0.000	10				
0.2	0.200	0.219	0.019	815	0.0 %	815				
2.5	2.500	2.563	0.063	9559	0.0 %	9559				
5.0	5.000	4.971	-0.029	18543	0.0 %	18543				
7.5	7.500	7.305	-0.195	27249	0.0 %	27249				
10.0	10.000	10.138	0.138	37817	0.0 %	37817				

LB93377 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Feb 2018 14:28:07

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	10	-	0.0000	7470A	PPB	19 Feb 2018 14:32:08	S	Std
0.2 - 1	50.2	815	-	0.2000	7470A	PPB	19 Feb 2018 14:34:16	S	Std
2.5 - 1	52.5	9559	-	2.5000	7470A	PPB	19 Feb 2018 14:36:25	S	Std
5.0 - 1	55	18543	-	5.0000	7470A	PPB	19 Feb 2018 14:38:33	S	Std
7.5 - 1	57.5	27249	-	7.5000	7470A	PPB	19 Feb 2018 14:40:41	S	Std
10.0 - 1	510	37817	-	10.0000	7470A	PPB	19 Feb 2018 14:46:53	S	Std
ICV06 - 1	ICV06	14075	3.7736	-	7470A	PPB	19 Feb 2018 14:49:26	U	SMPL
ICB06 - 1	ICB06	-107	-0.0281	-	7470A	PPB	19 Feb 2018 14:51:32	U	SMPL
CCV72 - 1	CCV72	18537	4.9698	-	7470A	PPB	19 Feb 2018 14:53:39	U	SMPL
CCB72 - 1	CCB72	-150	-0.0396	-	7470A	PPB	19 Feb 2018 14:55:45	U	SMPL
CRA - 1	CRA	774	0.2081	-	7470A	PPB	19 Feb 2018 14:57:52	U	SMPL
HighStd - 1	HighStd	37762	10.1233	-	7470A	PPB	19 Feb 2018 14:59:58	U	SMPL
ChkStd - 1	ChkStd	25363	6.7996	-	7470A	PPB	19 Feb 2018 15:02:05	U	SMPL
PB106639BL - 1	PBW	-226	-0.0600	-	7470A	PPB	19 Feb 2018 15:04:13	U	SMPL
PB106639BS - 1	LCSW	14505	3.8889	-	7470A	PPB	19 Feb 2018 15:06:21	U	SMPL
PB106639TB - 1	PB106600TB	-149	-0.0393	-	7470A	PPB	19 Feb 2018 15:08:29	U	SMPL
J1399-06 - 1	OR-02-021518-A	-61	-0.0158	-	7470A	PPB	19 Feb 2018 15:10:37	U	SMPL
J1399-06DUP - 1	OR-02-021518-ADUP	-23	-0.0056	-	7470A	PPB	19 Feb 2018 15:12:46	U	SMPL
J1399-06MS - 1	OR-02-021518-AMS	13066	3.5032	-	7470A	PPB	19 Feb 2018 15:14:54	U	SMPL
J1399-06MSD - 1	OR-02-021518-AMSD	13294	3.5643	-	7470A	PPB	19 Feb 2018 15:17:01	U	SMPL
CCV73 - 1	CCV73	18251	4.8931	-	7470A	PPB	19 Feb 2018 15:19:08	U	SMPL
CCB73 - 1	CCB73	-136	-0.0359	-	7470A	PPB	19 Feb 2018 15:21:15	U	SMPL
J1399-08 - 1	OR-02-021518-B	-72	-0.0187	-	7470A	PPB	19 Feb 2018 15:23:22	U	SMPL
J1401-02 - 1	SA-06-021518-A	-50	-0.0128	-	7470A	PPB	19 Feb 2018 15:25:29	U	SMPL
J1401-04 - 1	SA-06-021518-B	-19	-0.0045	-	7470A	PPB	19 Feb 2018 15:27:36	U	SMPL
J1575-06 - 1	SAMPLE-A	-21	-0.0050	-	7470A	PPB	19 Feb 2018 15:29:43	U	SMPL
J1575-12 - 1	SAMPLE-B	-21	-0.0050	-	7470A	PPB	19 Feb 2018 15:31:50	U	SMPL
PB106640BL - 1	PBW	-70	-0.0182	-	7470A	PPB	19 Feb 2018 15:33:58	U	SMPL
PB106640BS - 1	LCSW	13048	3.4983	-	7470A	PPB	19 Feb 2018 15:36:06	U	SMPL
J1577-01 - 1	COMPOSITE-A-TOTAL	-204	-0.0541	-	7470A	PPB	19 Feb 2018 15:38:14	U	SMPL
J1577-01DUP - 1	COMPOSITE-A-TOTALDUP	-91	-0.0238	-	7470A	PPB	19 Feb 2018 15:40:23	U	SMPL
J1577-01MS - 1	COMPOSITE-A-TOTALMS	14655	3.9291	-	7470A	PPB	19 Feb 2018 15:42:30	U	SMPL
CCV74 - 1	CCV74	18245	4.8915	-	7470A	PPB	19 Feb 2018 15:44:37	U	SMPL
CCB74 - 1	CCB74	-142	-0.0375	-	7470A	PPB	19 Feb 2018 15:46:44	U	SMPL
J1577-01MSD - 1	COMPOSITE-A-TOTALMSD	14469	3.8793	-	7470A	PPB	19 Feb 2018 15:48:52	U	SMPL
J1577-02 - 1	COMPOSITE-B-TOTAL	-322	-0.0857	-	7470A	PPB	19 Feb 2018 15:50:58	U	SMPL
J1577-03 - 1	COMPOSITE-A-DISSOLVED	-147	-0.0388	-	7470A	PPB	19 Feb 2018 15:53:06	U	SMPL
J1577-04 - 1	COMPOSITE-B-DISSOLVED	-180	-0.0477	-	7470A	PPB	19 Feb 2018 15:55:13	U	SMPL
J1601-02 - 1	DRUM-1	797	0.2142	-	7470A	PPB	19 Feb 2018 15:57:20	U	SMPL
J1399-06LX5 - 1		-77	-0.0200	-	7470A	PPB	19 Feb 2018 15:59:28	U	SMPL
J1399-06A - 1		13245	3.5511	-	7470A	PPB	19 Feb 2018 16:01:36	U	SMPL
J1577-01LX5 - 1		-121	-0.0318	-	7470A	PPB	19 Feb 2018 16:03:44	U	SMPL
J1577-01A - 1		15114	4.0522	-	7470A	PPB	19 Feb 2018 16:05:53	U	SMPL
CCV75 - 1	CCV75	18196	4.8783	-	7470A	PPB	19 Feb 2018 16:08:00	U	SMPL
CCB75 - 1	CCB75	-151	-0.0399	-	7470A	PPB	19 Feb 2018 16:10:08	U	SMPL