



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB93381

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	02/19/2018	09:35
2	J1565-01		1	>100.0	>212.0	02/19/2018	10:10
3	J1565-01DUP		1	>100.0	>212.0	02/19/2018	10:40
4	J1566-03		1	>100.0	>212.0	02/19/2018	11:05
5	J1566-05		1	>100.0	>212.0	02/19/2018	11:35
6	J1591-01		1	>100.0	>212.0	02/19/2018	11:55
7	J1591-02		1	>100.0	>212.0	02/19/2018	12:20
8	J1601-03		1	>100.0	>212.0	02/19/2018	12:45

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP21918

WorkList ID : 108678

Date : 2/19/2018 8:54:39 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/21/2018	Water	J1565-01	Flash Point	Cool 4 deg C	PSEG02	P62	18273	02/14/2018	1010A
02/21/2018	Water	J1566-03	Flash Point	Cool 4 deg C	PSEG01	P62	1452-56	02/14/2018	1010A
02/21/2018	Water	J1566-05	Flash Point	Cool 4 deg C	PSEG01	P62	1457-61	02/14/2018	1010A
03/02/2018	Water	J1591-01	Flash Point	Cool 4 deg C	DIST01	P41	MH-27	02/16/2018	1010A
03/02/2018	Water	J1591-02	Flash Point	Cool 4 deg C	DIST01	P41	MH-28	02/16/2018	1010A
02/21/2018	Water	J1601-03	Flash Point	Cool 4 deg C	PSEG01	P63	DRUM-1	02/16/2018	1010A

Date/Time

Received by:

Relinquished by:

Date/Time

Received by:

Relinquished by: