

Analysis Method:	9034	ANALYST:	Hiral	
Parameter:	Reactive Sulfide	SUPERVISOR REVIEW BY:	apatel	
Run Number:	LB93382	Constant:	16000	
		Normality1:	0.025	

Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE,0.025N,4LITRE	W2276
IODINE SOLUTION .025N 1L	W2283
Starch Solution, 4L	W2121

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB106654BL		1	50	50	2.00	0.00	1.86	1.86	0.14	0.00	0.00	02/19/2018	11:32
2	J1601-03		1	50	50	2.00	0.00	1.72	1.72	0.28	0.14	1.12	02/19/2018	11:35
3	J1601-03DUP		1	50	50	2.00	1.72	3.42	1.70	0.30	0.16	1.28	02/19/2018	11:38

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume