



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB93389

Temp.: 21.3 °C

BalanceID: WC SC-4

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/19/2018	15:35
2	CAL2	1	Water			7.00	02/19/2018	15:36
3	CAL3	1	Water			10.01	02/19/2018	15:37
4	ICV	1	Water			6.99	02/19/2018	15:40
5	CCV1	1	Water			2.01	02/19/2018	15:42
6	J1619-01	1	Solid	20.02	20	7.05	02/19/2018	15:44
7	J1619-01DUP	1	Solid	20.03	20	7.06	02/19/2018	15:45
8	J1619-02	1	Solid	20.03	20	6.49	02/19/2018	15:46
9	CCV2	1	Water			12.01	02/19/2018	15:47

1B43389

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY J1619 WorkList ID : 108718 Date : 2/19/2018 2:32:42 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/16/2018	Solid	J1619-01	Corrosivity	Cool 4 deg C	CLEA11	P41	BUS-WASH-SLUDGE-PIT	02/16/2018	9045C
02/16/2018	Solid	J1619-02	Corrosivity	Cool 4 deg C	CLEA11	P41	CHASSIS-WASH-SLUDGE-PI	02/16/2018	9045C

Date/Time 02-19-18 2:45 PM
Received by: JF
Relinquished by: _____

Date/Time 02-19-18 5:00 PM
Received by: CP
Relinquished by: JF