


7471B

LB93393
INSTRUMENT ID: CV1

Linear

 μ Abs.:

43081



A= 0.0000e+000

B= 2.3184e-004

C= -1.2034e-002

Rho= 0.9999871

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.022	-0.022	-44	0.000	-44				
0.2	0.200	0.195	-0.005	891	0.0 %	891				
2.5	2.500	2.517	0.017	10910	0.0 %	10910				
5.0	5.000	5.025	0.025	21725	0.0 %	21725				
7.5	7.500	7.510	0.010	32444	0.0 %	32444				
10	10.000	9.976	-0.024	43081	0.0 %	43081				

LB93393 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Feb 2018 10:31:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-44	-	0.0000	7471B	PPB	20 Feb 2018 10:33:02	S	Std
0.2 - 1	50.2	891	-	0.2000	7471B	PPB	20 Feb 2018 10:35:19	S	Std
2.5 - 1	52.5	10910	-	2.5000	7471B	PPB	20 Feb 2018 10:37:37	S	Std
5.0 - 1	55	21725	-	5.0000	7471B	PPB	20 Feb 2018 10:39:54	S	Std
7.5 - 1	57.5	32444	-	7.5000	7471B	PPB	20 Feb 2018 10:42:11	S	Std
10 - 1	510	43081	-	10.0000	7471B	PPB	20 Feb 2018 10:44:28	S	Std
ICV07 - 1	ICV07	17450	4.0336	-7471B	PPB	PPB	20 Feb 2018 10:48:17	U	SMPL
ICB07 - 1	ICB07	-177	-0.0531	-7471B	PPB	PPB	20 Feb 2018 10:50:33	U	SMPL
CCV76 - 1	CCV76	21813	5.0451	-7471B	PPB	PPB	20 Feb 2018 10:52:48	U	SMPL
CCB76 - 1	CCB76	-190	-0.0561	-7471B	PPB	PPB	20 Feb 2018 10:55:04	U	SMPL
CRA - 1	CRA	867	0.1890	-7471B	PPB	PPB	20 Feb 2018 10:57:21	U	SMPL
HighStd - 1	HighStd	43287	10.0236	-7471B	PPB	PPB	20 Feb 2018 10:59:37	U	SMPL
ChkStd - 1	ChkStd	29785	6.8933	-7471B	PPB	PPB	20 Feb 2018 11:01:53	U	SMPL
PB106680BL - 1	PBS	-289	-0.0790	-7471B	PPB	PPB	20 Feb 2018 11:04:11	U	SMPL
PB106680BS - 1	LCSS	14776	3.4136	-7471B	PPB	PPB	20 Feb 2018 11:06:29	U	SMPL
J1392-01 - 1	BU-03-021618	855	0.1862	-7471B	PPB	PPB	20 Feb 2018 11:08:47	U	SMPL
J1601-01 - 1	DRUM-2	808	0.1753	-7471B	PPB	PPB	20 Feb 2018 11:11:04	U	SMPL
J1605-01 - 1	EP-17	203	0.0350	-7471B	PPB	PPB	20 Feb 2018 11:13:21	U	SMPL
J1605-01DUP - 1	EP-17DUP	198	0.0339	-7471B	PPB	PPB	20 Feb 2018 11:15:39	U	SMPL
J1605-01MS - 1	EP-17MS	15924	3.6798	-7471B	PPB	PPB	20 Feb 2018 11:17:55	U	SMPL
CCV77 - 1	CCV77	21680	5.0143	-7471B	PPB	PPB	20 Feb 2018 11:20:11	U	SMPL
CCB77 - 1	CCB77	-223	-0.0637	-7471B	PPB	PPB	20 Feb 2018 11:22:27	U	SMPL
J1605-01MSD - 1	EP-17MSD	16183	3.7398	-7471B	PPB	PPB	20 Feb 2018 11:24:44	U	SMPL
J1605-02 - 1	EP-09	-17	-0.0160	-7471B	PPB	PPB	20 Feb 2018 11:26:59	U	SMPL
J1605-03 - 1	EP-01	231	0.0415	-7471B	PPB	PPB	20 Feb 2018 11:29:16	U	SMPL
J1606-01 - 1	BROOK	1496	0.3348	-7471B	PPB	PPB	20 Feb 2018 11:31:32	U	SMPL
J1605-01LX5 - 1		-26	-0.0181	-7471B	PPB	PPB	20 Feb 2018 11:33:49	U	SMPL
J1605-01A - 1		16852	3.8949	-7471B	PPB	PPB	20 Feb 2018 11:36:06	U	SMPL
CCV78 - 1	CCV78	21635	5.0038	-7471B	PPB	PPB	20 Feb 2018 11:38:23	U	SMPL
CCB78 - 1	CCB78	-221	-0.0633	-7471B	PPB	PPB	20 Feb 2018 11:40:41	U	SMPL