



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB93398

Temp.: 24.1 °C

SLOPE.: 99.0

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/20/2018	09:05
2	CAL2	1	Water			7.00	02/20/2018	09:07
3	CAL3	1	Water			10.01	02/20/2018	09:09
4	ICV	1	Water			7.00	02/20/2018	11:12
5	CCV1	1	Water			2.00	02/20/2018	11:14
6	J1621-01	1	Water			4.35	02/20/2018	11:16
7	J1621-01DUP	1	Water			4.36	02/20/2018	11:18
8	CCV2	1	Water			12.00	02/20/2018	11:20

WORKLIST(Hardcopy Internal Chain)

LB 93398

Worklist Name : pHJ1621

Worklist ID : 108735

Date : 2/20/2018 11:07:15 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/19/2017	Water	J1621-01	pH	Cool 4 deg C	STER09	P41	CELL-D-POOL-WATER(DE-IC	12/19/2017	SM 4500-PH B

Date/Time

02/20/18 11:10 AM

Received by:

SP

Relinquished by:

SP

Date/Time

02/20/18 11:25 AM

Received by:

SP

Relinquished by:

SP