


7470A

LB93403
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36983



A= 0.0000e+000

B= 2.6857e-004

C= 2.7206e-002

Rho= 0.9999086

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.014	0.014	-51	0.000	-51				
0.2	0.200	0.224	0.024	732	0.0 %	732				
2.5	2.500	2.410	-0.090	8874	0.0 %	8874				
5.0	5.000	5.035	0.035	18645	0.0 %	18645				
7.5	7.500	7.558	0.058	28040	0.0 %	28040				
10.0	10.000	9.960	-0.040	36983	0.0 %	36983				

LB93403 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Feb 2018 14:20:28

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	-51	-	0.00007470A	PPB	20 Feb 2018 14:23:32	S	Std
0.2 - 1	50.2	732	-	0.20007470A	PPB	20 Feb 2018 14:25:40	S	Std
2.5 - 1	52.5	8874	-	2.50007470A	PPB	20 Feb 2018 14:27:48	S	Std
5.0 - 1	55	18645	-	5.00007470A	PPB	20 Feb 2018 14:29:57	S	Std
7.5 - 1	57.5	28040	-	7.50007470A	PPB	20 Feb 2018 14:32:05	S	Std
10.0 - 1	511	36983	-	10.00007470A	PPB	20 Feb 2018 14:34:12	S	Std
ICV08 - 1	ICV08	14464	3.9118	-7470A	PPB	20 Feb 2018 14:36:53	U	SMPL
ICB08 - 1	ICB08	-167	-0.0176	-7470A	PPB	20 Feb 2018 14:38:59	U	SMPL
CCV79 - 1	CCV79	18615	5.0266	-7470A	PPB	20 Feb 2018 14:41:06	U	SMPL
CCB79 - 1	CCB79	-182	-0.0217	-7470A	PPB	20 Feb 2018 14:43:12	U	SMPL
CRA - 1	CRA	691	0.2128	-7470A	PPB	20 Feb 2018 14:45:20	U	SMPL
HighStd - 1	HighStd	37092	9.9889	-7470A	PPB	20 Feb 2018 14:47:26	U	SMPL
ChkStd - 1	ChkStd	26055	7.0248	-7470A	PPB	20 Feb 2018 14:49:33	U	SMPL
PB106672BL - 1	PBW	-240	-0.0373	-7470A	PPB	20 Feb 2018 14:51:42	U	SMPL
PB106672BS - 1	LCSW	14403	3.8954	-7470A	PPB	20 Feb 2018 14:53:50	U	SMPL
PB106672TB - 1	PB106634TB	-165	-0.0171	-7470A	PPB	20 Feb 2018 14:55:58	U	SMPL
J1392-02 - 1	BU-03-021618	-96	0.0014	-7470A	PPB	20 Feb 2018 14:58:06	U	SMPL
J1392-02DUP - 1	BU-03-021618DUP	-68	0.0089	-7470A	PPB	20 Feb 2018 15:00:14	U	SMPL
J1392-02MS - 1	BU-03-021618MS	13010	3.5213	-7470A	PPB	20 Feb 2018 15:02:23	U	SMPL
J1392-02MSD - 1	BU-03-021618MSD	12766	3.4558	-7470A	PPB	20 Feb 2018 15:04:30	U	SMPL
CCV80 - 1	CCV80	18817	5.0809	-7470A	PPB	20 Feb 2018 15:06:36	U	SMPL
CCB80 - 1	CCB80	-166	-0.0174	-7470A	PPB	20 Feb 2018 15:08:43	U	SMPL
J1604-03 - 1	TP-1	-23	0.0210	-7470A	PPB	20 Feb 2018 15:10:50	U	SMPL
J1604-06 - 1	TP-2	-2	0.0267	-7470A	PPB	20 Feb 2018 15:12:57	U	SMPL
J1604-09 - 1	TP-3	-118	-0.0045	-7470A	PPB	20 Feb 2018 15:15:03	U	SMPL
J1604-12 - 1	TP-4	-83	0.0049	-7470A	PPB	20 Feb 2018 15:17:10	U	SMPL
PB106673BL - 1	PBW	-160	-0.0158	-7470A	PPB	20 Feb 2018 15:19:18	U	SMPL
PB106673BS - 1	LCSW	13105	3.5468	-7470A	PPB	20 Feb 2018 15:21:25	U	SMPL
PB106673TB - 1	PB106636TB	-158	-0.0152	-7470A	PPB	20 Feb 2018 15:23:33	U	SMPL
J1601-03 - 1	DRUM-1	-101	0.0001	-7470A	PPB	20 Feb 2018 15:25:42	U	SMPL
J1601-03DUP - 1	DRUM-1DUP	-87	0.0038	-7470A	PPB	20 Feb 2018 15:27:51	U	SMPL
J1601-03MS - 1	DRUM-1MS	13599	3.6795	-7470A	PPB	20 Feb 2018 15:29:57	U	SMPL
CCV81 - 1	CCV81	18902	5.1037	-7470A	PPB	20 Feb 2018 15:32:04	U	SMPL
CCB81 - 1	CCB81	-158	-0.0152	-7470A	PPB	20 Feb 2018 15:34:11	U	SMPL
J1601-03MSD - 1	DRUM-1MSD	13043	3.5301	-7470A	PPB	20 Feb 2018 15:36:19	U	SMPL
J1392-02LX5 - 1		-162	-0.0163	-7470A	PPB	20 Feb 2018 15:38:26	U	SMPL
J1392-02A - 1		13762	3.7232	-7470A	PPB	20 Feb 2018 15:40:33	U	SMPL
J1601-03LX5 - 1		-162	-0.0163	-7470A	PPB	20 Feb 2018 15:42:40	U	SMPL
J1601-03A - 1		13652	3.6937	-7470A	PPB	20 Feb 2018 15:44:48	U	SMPL
CCV82 - 1	CCV82	18683	5.0449	-7470A	PPB	20 Feb 2018 15:46:55	U	SMPL
CCB82 - 1	CCB82	-167	-0.0176	-7470A	PPB	20 Feb 2018 15:49:03	U	SMPL