


245.1

LB93405
INSTRUMENT ID: CV1

Linear

 μ Abs.:

34731



A= 0.0000e+000

B= 2.8871e-004

C= -2.3352e-002

Rho= 0.9999147

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.035	-0.035	-41	0.000	-41				
0.05	0.050	0.044	-0.006	235	0.0 %	235				
0.2	0.200	0.186	-0.014	725	0.0 %	725				
2.5	2.500	2.545	0.045	8896	0.0 %	8896				
5.0	5.000	5.083	0.083	17688	0.0 %	17688				
7.5	7.500	7.422	-0.078	25790	0.0 %	25790				
10.0	10.000	10.004	0.004	34731	0.0 %	34731				

LB93405 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 20 Feb 2018 16:07:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-41	-	0.0000	245.1	PPB	20 Feb 2018 16:10:26	S	Std
0.05 - 1	50.5	235	-	0.0500	245.1	PPB	20 Feb 2018 16:12:34	S	Std
0.2 - 1	50.5	725	-	0.2000	245.1	PPB	20 Feb 2018 16:14:43	S	Std
2.5 - 1	55	8896	-	2.5000	245.1	PPB	20 Feb 2018 16:16:51	S	Std
5.0 - 1	57.5	17688	-	5.0000	245.1	PPB	20 Feb 2018 16:18:59	S	Std
7.5 - 1	510	25790	-	7.5000	245.1	PPB	20 Feb 2018 16:21:06	S	Std
10.0 - 1		34731	-	10.0000	245.1	PPB	20 Feb 2018 16:25:46	S	Std
ICV09 - 1	ICV09	13945	4.0027	-245.1		PPB	20 Feb 2018 16:28:31	U	SMPL
ICB09 - 1	ICB09	-135	-0.0623	-245.1		PPB	20 Feb 2018 16:30:37	U	SMPL
CCV83 - 1	CCV83	17899	5.1443	-245.1		PPB	20 Feb 2018 16:32:44	U	SMPL
CCB83 - 1	CCB83	-166	-0.0713	-245.1		PPB	20 Feb 2018 16:34:50	U	SMPL
CRA - 1	CRA	253	0.0497	-245.1		PPB	20 Feb 2018 16:43:53	U	SMPL
HighStd - 1	HighStd	34474	9.9297	-245.1		PPB	20 Feb 2018 16:46:00	U	SMPL
ChkStd - 1	ChkStd	24374	7.0137	-245.1		PPB	20 Feb 2018 16:48:07	U	SMPL
PB106674BL - 1	PBW	-88	-0.0488	-245.1		PPB	20 Feb 2018 16:52:40	U	SMPL
PB106674BS - 1	LCSW	12917	3.7059	-245.1		PPB	20 Feb 2018 16:54:48	U	SMPL
J1591-01 - 1	MH-27	-124	-0.0592	-245.1		PPB	20 Feb 2018 16:56:56	U	SMPL
J1591-02 - 1	MH-28	-240	-0.0926	-245.1		PPB	20 Feb 2018 16:59:04	U	SMPL
J1611-01 - 1	APT	-38	-0.0343	-245.1		PPB	20 Feb 2018 17:01:13	U	SMPL
J1611-01DUP - 1	APTDUP	-48	-0.0372	-245.1		PPB	20 Feb 2018 17:03:19	U	SMPL
CCV84 - 1	CCV84	17687	5.0831	-245.1		PPB	20 Feb 2018 17:05:26	U	SMPL
CCB84 - 1	CCB84	-146	-0.0655	-245.1		PPB	20 Feb 2018 17:07:33	U	SMPL
J1611-01MS - 1	APTMS	14618	4.1970	-245.1		PPB	20 Feb 2018 17:09:41	U	SMPL
J1611-01MSD - 1	APTMSD	14591	4.1892	-245.1		PPB	20 Feb 2018 17:11:47	U	SMPL
J1611-02 - 1	BASEMENT	-361	-0.1276	-245.1		PPB	20 Feb 2018 17:13:54	U	SMPL
J1611-01LX5 - 1		-79	-0.0462	-245.1		PPB	20 Feb 2018 17:16:00	U	SMPL
J1611-01A - 1		15325	4.4011	-245.1		PPB	20 Feb 2018 17:18:08	U	SMPL
CCV85 - 1	CCV85	17782	5.1105	-245.1		PPB	20 Feb 2018 17:20:16	U	SMPL
CCB85 - 1	CCB85	-166	-0.0713	-245.1		PPB	20 Feb 2018 17:22:24	U	SMPL