



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB93407

Temp.: 22.6 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 2.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/20/2018	17:30
2	CAL2	1	Water			7.01	02/20/2018	17:31
3	CAL3	1	Water			10.03	02/20/2018	17:32
4	ICV	1	Water			7.00	02/20/2018	17:33
5	CCV1	1	Water			2.02	02/20/2018	17:35
6	J1391-02	1	Solid	20.02	20	5.83	02/20/2018	17:36
7	J1391-02DUP	1	Solid	20.04	20	5.84	02/20/2018	17:37
8	J1391-04	1	Solid	20.02	20	5.73	02/20/2018	17:39
9	J1398-04	1	Solid	20.01	20	8.30	02/20/2018	17:41
10	J1403-06	1	Solid	20.04	20	7.72	02/20/2018	17:42
11	J1404-06	1	Solid	20.05	20	6.33	02/20/2018	17:43
12	J1622-02	1	Solid	20.03	20	7.29	02/20/2018	17:45
13	J1623-04	1	Solid	20.02	20	8.70	02/20/2018	17:46
14	J1623-08	1	Solid	20.02	20	7.83	02/20/2018	17:48
15	J1623-12	1	Solid	20.01	20	7.65	02/20/2018	17:50
16	CCV2	1	Water			12.02	02/20/2018	17:51
17	J1623-16	1	Solid	20.03	20	8.06	02/20/2018	18:00
18	J1623-20	1	Solid	20.04	20	7.95	02/20/2018	18:01
19	J1624-02	1	Solid	20.02	20	6.92	02/20/2018	18:03
20	J1624-04	1	Solid	20.05	20	7.84	02/20/2018	18:05

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	J1624-06	1	Solid	20.03	20	7.36	02/20/2018	18:06
22	CCV3	1	Water			2.01	02/20/2018	18:09

1393407

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosion j1624

WorkList ID : 108750

Date : 2/20/2018 3:24:05 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/20/2018	Solid	J1391-02	Corrosivity	Cool 4 deg C	PSEG05	P61	AU-05-022018-A	02/20/2018	9045C
02/20/2018	Solid	J1391-04	Corrosivity	Cool 4 deg C	PSEG05	P61	AU-05-022018-B	02/20/2018	9045C
02/20/2018	Solid	J1398-04	Corrosivity	Cool 4 deg C	PSEG05		OK-02-022018	02/20/2018	9045C
02/20/2018	Solid	J1403-06	Corrosivity	Cool 4 deg C	PSEG05		SU-04-022018	02/20/2018	9045C
02/20/2018	Solid	J1404-06	Corrosivity	Cool 4 deg C	PSEG05		TR-05-022018	02/20/2018	9045C
02/20/2018	Solid	J1622-02	Corrosivity	Cool 4 deg C	PSEG01	P63	MAD-SUB-TRENCH	02/20/2018	9045C
02/20/2018	Solid	J1623-04	Corrosivity	Cool 4 deg C	PSEG01	P63	TP-1-2	02/20/2018	9045C
02/20/2018	Solid	J1623-08	Corrosivity	Cool 4 deg C	PSEG01	P63	TP-3-4	02/20/2018	9045C
02/20/2018	Solid	J1623-12	Corrosivity	Cool 4 deg C	PSEG01	P63	TP-5-6	02/20/2018	9045C
02/20/2018	Solid	J1623-16	Corrosivity	Cool 4 deg C	PSEG01	P63	TP-7-8	02/20/2018	9045C
02/20/2018	Solid	J1623-20	Corrosivity	Cool 4 deg C	PSEG01	P63	TP-9-10	02/20/2018	9045C
02/20/2018	Solid	J1624-02	Corrosivity	Cool 4 deg C	PSEG01	P63	TP-1	02/20/2018	9045C
02/20/2018	Solid	J1624-04	Corrosivity	Cool 4 deg C	PSEG01	P63	TP-2	02/20/2018	9045C
02/20/2018	Solid	J1624-06	Corrosivity	Cool 4 deg C	PSEG01	P63	TP-3	02/20/2018	9045C

Date/Time 02-20-17 3:30 PM

Received by: JA
Relinquished by: CS

Date/Time 02-20-17 6:25 PM

Received by: CS
Relinquished by: JA